

LABORATORY REPORT

If you have any questions concerning this report, please do not hesitate to call us at (800) 332-4345 or (574) 233-4777.

This report may not be reproduced, except in full, without written approval from EEA.

STATE CERTIFICATION LIST

State	Certification	State	Certification
Alabama	40700	Missouri	880
Alaska	IN00035	Montana	CERT0026
Arizona	AZ0432	Nebraska	NE-OS-05-04
Arkansas	IN00035	Nevada	IN00035
California	2920	New Hampshire*	2124
Colorado	IN00035	New Jersey*	IN598
Colorado Radiochemistry	IN00035	New Mexico	IN00035
Connecticut	PH-0132	New York*	11398
Delaware	IN035	North Carolina	18700
Florida(Primary AB)*	E87775	North Dakota	R-035
Georgia	929	Ohio	87775
Hawaii	IN035	Oklahoma	D9508
Idaho	IN00035	Oregon*	4156
Illinois*	200001	Pennsylvania*	68-00466
Illinois Microbiology	17767	Puerto Rico	IN00035
Illinois Radiochemistry	IN00035	Rhode Island	LAO00343
Indiana Chemistry	C-71-01	South Carolina	95005
Indiana Microbiology	M-76-07	South Dakota	IN00035
Iowa	098	Tennessee	TN02973
Kansas*	E-10233	Texas*	T104704187
Kentucky	90056	Texas/TCEQ	TX207
Louisiana*	LA014	Utah*	IN00035
Maine	IN00035	Vermont	VT-8775
Maryland	209	Virginia*	460275
Massachusetts	M-IN035	Washington	C837
Michigan	9926	West Virginia	9927 C
Minnesota*	018-999-338	Wisconsin	999766900
Mississippi	IN035	Wyoming	IN035
EPA	IN00035		

*NELAP/TNI Recognized Accreditation Bodies

NELAC NARRATIVE PAGE

Client: Leavenworth Waterworks Dept.

Report #: 508807NP

Eurofins Eaton Analytical, LLC is a NELAP accredited laboratory. All reported results meet the requirements of the NELAC standards, unless otherwise noted.

EEA contact person: Caleb Hunsberger

NELAP requires complete reporting of deviations from method requirements, regardless of the suspected impact on the data. Quality control failures not reported within the report summary are noted here.

There were no quality control failures.

Note: This report may not be reproduced, except in full, without written approval from EEA. EEA is accredited by the National Environmental Laboratory Accreditation Program (NELAP).

		01/30/2021
Authorized Signature	Title	Date

110 South Hill Street
 South Bend, IN 46617
 Tel: (574) 233-4777
 Fax: (574) 233-8207
 1 800 332 4345

Laboratory Report

Client: Leavenworth Waterworks Dept.
 Attn: Joel Mahnken
 PO Box 576
 Leavenworth, KS 66048

Report: 508807
 Priority: Standard Written
 Status: Final
 PWS ID: KS2010317
 Lab ID#: E-10233
 Program Code: PE

Sample Information					
EEA ID #	Client ID	Method	Collected Date / Time	Collected By:	Received Date / Time
4820255	Plant 1 Tap-LV Waterworks	537.1	01/25/21 14:06	Client	01/26/21 08:30
4820256	Plant 2 Tap-LV Waterworks	537.1	01/25/21 14:38	Client	01/26/21 08:30

Report Summary

Detailed quantitative results are presented on the following pages. The results presented relate only to the samples provided for analysis.

We appreciate the opportunity to provide you with this analysis. If you have any questions concerning this report, please do not hesitate to call Caleb Hunsberger at (574) 233-4777.

Note: This report may not be reproduced, except in full, without written approval from EEA. EEA is accredited by the National Environmental Laboratory Accreditation Program (NELAP).

Caleb Hunsberger ASM

Authorized Signature

Title

01/30/2021

Date

Client Name: Leavenworth Waterworks Dept.
 Report #: 508807

Sampling Point: Plant 1 Tap-LV Waterworks

PWS ID: KS2010317

EEA Methods									
Analyte ID #	Analyte	Method	Reg Limit	MRL†	Result	Units	Preparation Date	Analyzed Date	EEA ID #
335-67-1	Perfluorooctanoic acid (PFOA)	537.1	---	2.0	< 2.0	ng/L	01/28/21 07:57	01/29/21 02:28	4820255
1763-23-1	Perfluorooctanesulfonic acid (PFOS)	537.1	---	2.0	< 2.0	ng/L	01/28/21 07:57	01/29/21 02:28	4820255
375-73-5	Perfluorobutanesulfonic acid (PFBS)	537.1	---	2.0	< 2.0	ng/L	01/28/21 07:57	01/29/21 02:28	4820255
375-85-9	Perfluoroheptanoic acid (PFHpA)	537.1	---	2.0	< 2.0	ng/L	01/28/21 07:57	01/29/21 02:28	4820255
355-46-4	Perfluorohexanesulfonic acid (PFHxS)	537.1	---	2.0	< 2.0	ng/L	01/28/21 07:57	01/29/21 02:28	4820255
375-95-1	Perfluorononanoic acid (PFNA)	537.1	---	2.0	< 2.0	ng/L	01/28/21 07:57	01/29/21 02:28	4820255
335-76-2	Perfluorodecanoic acid (PFDA)	537.1	---	2.0	< 2.0	ng/L	01/28/21 07:57	01/29/21 02:28	4820255
307-24-4	Perfluorohexanoic acid (PFHxA)	537.1	---	2.0	< 2.0	ng/L	01/28/21 07:57	01/29/21 02:28	4820255
307-55-1	Perfluorododecanoic acid (PFDoA)	537.1	---	2.0	< 2.0	ng/L	01/28/21 07:57	01/29/21 02:28	4820255
72629-94-8	Perfluorotridecanoic acid (PFTrDA)	537.1	---	2.0	< 2.0	ng/L	01/28/21 07:57	01/29/21 02:28	4820255
2058-94-8	Perfluoroundecanoic acid (PFUnA)	537.1	---	2.0	< 2.0	ng/L	01/28/21 07:57	01/29/21 02:28	4820255
2991-50-6	N-ethyl Perfluorooctanesulfonamidoacetic acid	537.1	---	2.0	< 2.0	ng/L	01/28/21 07:57	01/29/21 02:28	4820255
2355-31-9	N-methyl Perfluorooctanesulfonamidoacetic acid	537.1	---	2.0	< 2.0	ng/L	01/28/21 07:57	01/29/21 02:28	4820255
13252-13-6	HFPO-DA/GenX	537.1	---	2.0	< 2.0	ng/L	01/28/21 07:57	01/29/21 02:28	4820255
919005-14-4	ADONA	537.1	---	2.0	< 2.0	ng/L	01/28/21 07:57	01/29/21 02:28	4820255
756426-58-1	9CI-PF3ONS/F-53B Major	537.1	---	2.0	< 2.0	ng/L	01/28/21 07:57	01/29/21 02:28	4820255
763051-92-9	11CI-PF3OUdS/F-53B Minor	537.1	---	2.0	< 2.0	ng/L	01/28/21 07:57	01/29/21 02:28	4820255
376-06-7	Perfluorotetradecanoic acid (PFTeDA)	537.1	---	2.0	< 2.0	ng/L	01/28/21 07:57	01/29/21 02:28	4820255

Sampling Point: Plant 2 Tap-LV Waterworks

PWS ID: KS2010317

EEA Methods									
Analyte ID #	Analyte	Method	Reg Limit	MRL†	Result	Units	Preparation Date	Analyzed Date	EEA ID #
335-67-1	Perfluorooctanoic acid (PFOA)	537.1	---	2.0	< 2.0	ng/L	01/28/21 07:57	01/29/21 02:39	4820256
1763-23-1	Perfluorooctanesulfonic acid (PFOS)	537.1	---	2.0	< 2.0	ng/L	01/28/21 07:57	01/29/21 02:39	4820256
375-73-5	Perfluorobutanesulfonic acid (PFBS)	537.1	---	2.0	< 2.0	ng/L	01/28/21 07:57	01/29/21 02:39	4820256
375-85-9	Perfluoroheptanoic acid (PFHpA)	537.1	---	2.0	< 2.0	ng/L	01/28/21 07:57	01/29/21 02:39	4820256
355-46-4	Perfluorohexanesulfonic acid (PFHxS)	537.1	---	2.0	< 2.0	ng/L	01/28/21 07:57	01/29/21 02:39	4820256
375-95-1	Perfluorononanoic acid (PFNA)	537.1	---	2.0	< 2.0	ng/L	01/28/21 07:57	01/29/21 02:39	4820256
335-76-2	Perfluorodecanoic acid (PFDA)	537.1	---	2.0	< 2.0	ng/L	01/28/21 07:57	01/29/21 02:39	4820256
307-24-4	Perfluorohexanoic acid (PFHxA)	537.1	---	2.0	< 2.0	ng/L	01/28/21 07:57	01/29/21 02:39	4820256
307-55-1	Perfluorododecanoic acid (PFDoA)	537.1	---	2.0	< 2.0	ng/L	01/28/21 07:57	01/29/21 02:39	4820256
72629-94-8	Perfluorotridecanoic acid (PFTrDA)	537.1	---	2.0	< 2.0	ng/L	01/28/21 07:57	01/29/21 02:39	4820256
2058-94-8	Perfluoroundecanoic acid (PFUnA)	537.1	---	2.0	< 2.0	ng/L	01/28/21 07:57	01/29/21 02:39	4820256
2991-50-6	N-ethyl Perfluorooctanesulfonamidoacetic acid	537.1	---	2.0	< 2.0	ng/L	01/28/21 07:57	01/29/21 02:39	4820256
2355-31-9	N-methyl Perfluorooctanesulfonamidoacetic acid	537.1	---	2.0	< 2.0	ng/L	01/28/21 07:57	01/29/21 02:39	4820256
13252-13-6	HFPO-DA/GenX	537.1	---	2.0	< 2.0	ng/L	01/28/21 07:57	01/29/21 02:39	4820256
919005-14-4	ADONA	537.1	---	2.0	< 2.0	ng/L	01/28/21 07:57	01/29/21 02:39	4820256
756426-58-1	9CI-PF3ONS/F-53B Major	537.1	---	2.0	< 2.0	ng/L	01/28/21 07:57	01/29/21 02:39	4820256
763051-92-9	11CI-PF3OUdS/F-53B Minor	537.1	---	2.0	< 2.0	ng/L	01/28/21 07:57	01/29/21 02:39	4820256
376-06-7	Perfluorotetradecanoic acid (PFTeDA)	537.1	---	2.0	< 2.0	ng/L	01/28/21 07:57	01/29/21 02:39	4820256

† EEA has demonstrated it can achieve these report limits in reagent water, but can not document them in all sample matrices.

Reg Limit Type:	MCL	SMCL	AL
Symbol:	*	^	!

Lab Definitions

Continuing Calibration Check Standard (CCC) / Continuing Calibration Verification (CCV) / Initial Calibration Verification Standard (ICV) / Initial Performance Check (IPC) - is a standard containing one or more of the target analytes that is prepared from the same standards used to calibrate the instrument. This standard is used to verify the calibration curve at the beginning of each analytical sequence, and may also be analyzed throughout and at the end of the sequence. The concentration of continuing standards may be varied, when prescribed by the reference method, so that the range of the calibration curve is verified on a regular basis. CCL, CCM, and CCH are the CCC standards at low, mid, and high concentration levels, respectively.

Internal Standards (IS) - are pure compounds with properties similar to the analytes of interest, which are added to field samples or extracts, calibration standards, and quality control standards at a known concentration. They are used to measure the relative responses of the analytes of interest and surrogates in the sample, calibration standard or quality control standard.

Laboratory Duplicate (LD) - is a field sample aliquot taken from the same sample container in the laboratory and analyzed separately using identical procedures. Analysis of laboratory duplicates provides a measure of the precision of the laboratory procedures.

Laboratory Fortified Blank (LFB) / Laboratory Control Sample (LCS) - is an aliquot of reagent water to which known concentrations of the analytes of interest are added. The LFB is analyzed exactly the same as the field samples. LFBs are used to determine whether the method is in control. FBL, FBM, and FBH are the LFB samples at low, mid, and high concentration levels, respectively.

Laboratory Method Blank (LMB) / Laboratory Reagent Blank (LRB) - is a sample of reagent water included in the sample batch analyzed in the same way as the associated field samples. The LMB is used to determine if method analytes or other background contamination have been introduced during the preparation or analytical procedure. The LMB is analyzed exactly the same as the field samples.

Laboratory Trip Blank (LTB) / Field Reagent Blank (FRB) - is a sample of laboratory reagent water placed in a sample container in the laboratory and treated as a field sample, including storage, preservation, and all analytical procedures. The FRB/LTB container follows the collection bottles to and from the collection site, but the FRB/LTB is not opened at any time during the trip. The FRB/LTB is primarily a travel blank used to verify that the samples were not contaminated during shipment.

If applicable, the calculation of the matrix spike (MS) or matrix spike duplicate (MSD) percent recovery is as follows: $(MS \text{ or } MSD \text{ value} - \text{Sample value}) * 100 / \text{spike target} / \text{dilution factor} = \text{Recovery } \%$

Matrix Spike Duplicate Sample (MSD) / Laboratory Fortified Sample Matrix Duplicate (LFSMD) - is a sample aliquot taken from the same field sample source as the Matrix Spike Sample to which known quantities of the analytes of interest are added in the laboratory. The MSD is analyzed exactly the same as the field samples. Analysis of the MSD provides a measure of the precision of the laboratory procedures in a specific matrix. SDL, SDM, and SDH / LFSMDL, LFSMDM, and LFSMDH are the MSD or LFSMD at low, mid, and high concentration levels, respectively.

Matrix Spike Sample (MS) / Laboratory Fortified Sample Matrix (LFSM) - is a sample aliquot taken from field sample source to which known quantities of the analytes of interest are added in the laboratory. The MS is analyzed exactly the same as the field samples. The purpose is to demonstrate recovery of the analytes from a sample matrix to determine if the specific matrix contributes bias to the analytical results. MSL, MSM, and MSH / LFSML, LFSMM, and LFSMH are the MS or LFSM at low, mid, and high concentration levels, respectively.

Quality Control Standard (QCS) / Second Source Calibration Verification (SSCV) - is a solution containing known concentrations of the analytes of interest prepared from a source different from the source of the calibration standards. The solution is obtained from a second manufacturer or lot if the lot can be demonstrated by the manufacturer as prepared independently from other lots. The QCS sample is analyzed using the same procedures as field samples. The QCS is used as a check on the calibration standards used in the method on a routine basis.

Reporting Limit Check (RLC) / Initial Calibration Check Standard (ICCS) - is a procedural standard that is analyzed each day to evaluate instrument performance at or below the minimum reporting limit (MRL).

Surrogate Standard (SS) / Surrogate Analyte (SUR) - is a pure compound with properties similar to the analytes of interest, which is highly unlikely to be found in any field sample, that is added to the field samples, calibration standards, blanks and quality control standards before sample preparation. The SS is used to evaluate the efficiency of the sample preparation process.



Eaton Analytical

110 S. Hill Street
South Bend, IN 46617
T: 1.800.332.4345
F: 1.574.233.8207

Order # 416689
Batch # 508807

www.EurofinsUS.com/Eaton

CHAIN OF CUSTODY RECORD

Shaded area for EEA use only

Page 1 of 1

REPORT TO:		SAMPLER (Signature)		COMPLIANCE MONITORING		PWS ID #		STATE (sample origin)		PROJECT NAME		PO#		MATRIX CODE		TURNAROUND TIME	
BILL TO:		Yes		No		K52010317		KS		Preservative Checks							
LAB Number		DATE		TIME		AM		PM		SAMPLING SITE		TEST NAME		Residual Chlorine (P/A)		CHLORINATED	
		DATE		TIME		AM		PM				YES		NO		# OF CONTAINERS	
1	4820255	1/25/21	14:00			✓				Plant #1 Tap - LV Waterworks							2
2	↓ 256	1/25/21	14:30			✓				Plant #2 Tap - LV Waterworks							2
3																	
4																	
5																	
6																	
7																	
8																	
9																	
10																	
11																	
12																	
13																	
14																	

LAB RESERVES THE RIGHT TO RETURN UNUSED PORTIONS OF NON-AQUEOUS SAMPLES TO CLIENT

RELINQUISHED BY: (Signature) [Signature] DATE 1-25-21 TIME 14:50 AM | PM

RELINQUISHED BY: (Signature) [Signature] DATE _____ TIME _____ AM | PM

RECEIVED BY: (Signature) [Signature] DATE 1-26-2020 TIME 0830 AM | PM

RECEIVED FOR LABORATORY BY: [Signature] DATE _____ TIME _____ AM | PM

LAB COMMENTS: _____

CONDITIONS UPON RECEIPT (check one):
 () Cold: Wet/Blue _____ Ambient _____ °C Upon Receipt: _____ N/A

MATRIX CODES:
 DW-DRINKING WATER RW-REAGENT WATER GW-
 GROUND WATER EW-EXPOSURE WATER SW-
 SURFACE WATER PW-POOL WATER
 WW-WASTE WATER

SW = Standard Written: (15 working days) 0% RV = Rush Verbal: (5 working days)
 50% RW = Rush Written: (5 working days) 75%
 * Please call, expedited service not available for all testing

100%
 125%
 CALL
 CALL

STAT* = Less than 48 hours

100%
 125%
 CALL
 CALL

Samples received unannounced with less than 48 hours holding time remaining may be subject to additional charges.
 06-LO-F0435 Issue 8.0 Effective Date: 2020-05-15

Eurofins Eaton Analytical

Run Log

Run ID: 284752 Method: 537.1

<u>Type</u>	<u>Sample Id</u>	<u>Sample Site</u>	<u>Matrix</u>	<u>Instrument ID</u>	<u>Analysis Date</u>	<u>Calibration File</u>
CCL	4819836		OS	FL	01/26/2021 07:04	012621M537_1a-FL.mdb
LRB	4819731		RW	FL	01/26/2021 07:30	012621M537_1a-FL.mdb
FBL	4819732		RW	FL	01/26/2021 07:43	012621M537_1a-FL.mdb
FBM	4819735		RW	FL	01/26/2021 07:56	012621M537_1a-FL.mdb
CCM	4819837		OS	FL	01/26/2021 11:17	012621M537_1a-FL.mdb
CCH	4819841		OS	FL	01/26/2021 15:01	012621M537_1a-FL.mdb
CCM	4820887		OS	FL	01/26/2021 17:24	012621M537_1a-FL.mdb
CCH	4820890		OS	FL	01/27/2021 09:21	012621M537_1a-FL.mdb
CCM	4820888		OS	FL	01/27/2021 09:49	012621M537_1a-FL.mdb

QC Summary Report

Sample Type	Analyte	Method	MDA95	Client ID	Result Flag	Amount	Target	Units	% Recovery	Recovery Limits	RPD	RPD Limit	Dil Factor	Extracted	Analyzed	EEA ID #
CCL	Perfluorooctanoic acid (PFOA)	537.1	2.0	--		1.6661	2.0	ng/L	83	50 - 150	---	---	1.0	01/22/2021 08:59	01/26/2021 07:04	4819836
CCL	Perfluorooctanesulfonic acid (PFOS)	537.1	2.0	--		1.8044	2.0	ng/L	90	50 - 150	---	---	1.0	01/22/2021 08:59	01/26/2021 07:04	4819836
CCL	IS-NMeFOSAA-d3	537.1	N/A	--		207534	207534	ng/L	100	50 - 150	---	---	1.0	01/22/2021 08:59	01/26/2021 07:04	4819836
CCL	IS-PFOA-13C2	537.1	N/A	--		539839	539839	ng/L	100	50 - 150	---	---	1.0	01/22/2021 08:59	01/26/2021 07:04	4819836
CCL	IS-PFOS-13C4	537.1	N/A	--		271100	271100	ng/L	100	50 - 150	---	---	1.0	01/22/2021 08:59	01/26/2021 07:04	4819836
CCL	SS-NEFOSAA-d5	537.1	N/A	--		162.8880	160	ng/L	102	70 - 130	---	---	1.0	01/22/2021 08:59	01/26/2021 07:04	4819836
CCL	SS-PFDA-13C2	537.1	N/A	--		38.5257	40.0	ng/L	96	70 - 130	---	---	1.0	01/22/2021 08:59	01/26/2021 07:04	4819836
CCL	SS-PFHXA-13C2	537.1	N/A	--		40.0764	40.0	ng/L	100	70 - 130	---	---	1.0	01/22/2021 08:59	01/26/2021 07:04	4819836
CCL	Perfluorobutanesulfonic acid (PFBS)	537.1	2.0	--		1.5050	2.0	ng/L	75	50 - 150	---	---	1.0	01/22/2021 08:59	01/26/2021 07:04	4819836
CCL	Perfluorheptanoic acid (PFHpA)	537.1	2.0	--		1.5456	2.0	ng/L	77	50 - 150	---	---	1.0	01/22/2021 08:59	01/26/2021 07:04	4819836
CCL	Perfluorohexanesulfonic acid (PFHxS)	537.1	2.0	--		1.4691	2.0	ng/L	73	50 - 150	---	---	1.0	01/22/2021 08:59	01/26/2021 07:04	4819836
CCL	Perfluorononanoic acid (PFNA)	537.1	2.0	--		1.6915	2.0	ng/L	85	50 - 150	---	---	1.0	01/22/2021 08:59	01/26/2021 07:04	4819836
CCL	Perfluorodecanoic acid (PFDA)	537.1	2.0	--		1.6594	2.0	ng/L	83	50 - 150	---	---	1.0	01/22/2021 08:59	01/26/2021 07:04	4819836
CCL	Perfluorohexanoic acid (PFHxA)	537.1	2.0	--		1.6475	2.0	ng/L	82	50 - 150	---	---	1.0	01/22/2021 08:59	01/26/2021 07:04	4819836
CCL	Perfluorododecanoic acid (PFDoA)	537.1	2.0	--		1.7017	2.0	ng/L	85	50 - 150	---	---	1.0	01/22/2021 08:59	01/26/2021 07:04	4819836
CCL	Perfluorotridecanoic acid (PFTDA)	537.1	2.0	--		1.7493	2.0	ng/L	87	50 - 150	---	---	1.0	01/22/2021 08:59	01/26/2021 07:04	4819836
CCL	Perfluoroundecanoic acid (PFUnA)	537.1	2.0	--		1.6026	2.0	ng/L	80	50 - 150	---	---	1.0	01/22/2021 08:59	01/26/2021 07:04	4819836
CCL	N-ethyl Perfluorooctanesulfonamidoacetic acid	537.1	2.0	--		1.8482	2.0	ng/L	92	50 - 150	---	---	1.0	01/22/2021 08:59	01/26/2021 07:04	4819836
CCL	N-methyl Perfluorooctanesulfonamidoacetic acid	537.1	2.0	--		1.8774	2.0	ng/L	94	50 - 150	---	---	1.0	01/22/2021 08:59	01/26/2021 07:04	4819836
CCL	HFPO-DA/GenX	537.1	2.0	--		1.5492	2.0	ng/L	77	50 - 150	---	---	1.0	01/22/2021 08:59	01/26/2021 07:04	4819836
CCL	ADONA	537.1	2.0	--		1.3450	2.0	ng/L	67	50 - 150	---	---	1.0	01/22/2021 08:59	01/26/2021 07:04	4819836
CCL	9CI-PF3ONS/F-53B Major	537.1	2.0	--		1.8330	2.0	ng/L	92	50 - 150	---	---	1.0	01/22/2021 08:59	01/26/2021 07:04	4819836
CCL	11CI-PF3OUdS/F-53B Minor	537.1	2.0	--		1.8823	2.0	ng/L	94	50 - 150	---	---	1.0	01/22/2021 08:59	01/26/2021 07:04	4819836
CCL	Perfluorotetradecanoic acid (PFTeDA)	537.1	2.0	--		1.8441	2.0	ng/L	92	50 - 150	---	---	1.0	01/22/2021 08:59	01/26/2021 07:04	4819836
CCL	SS-HFO-DA-13C3	537.1	N/A	--		39.9852	40.0	ng/L	100	70 - 130	---	---	1.0	01/22/2021 08:59	01/26/2021 07:04	4819836
LRB	Perfluorooctanoic acid (PFOA)	537.1	2.0	--	<	2.0		ng/L	---	---	---	---	1.0	01/25/2021 07:20	01/26/2021 07:30	4819731
LRB	Perfluorooctanesulfonic acid (PFOS)	537.1	2.0	--	<	2.0		ng/L	---	---	---	---	1.0	01/25/2021 07:20	01/26/2021 07:30	4819731
LRB	IS-NMeFOSAA-d3	537.1	N/A	--		213441	207534	ng/L	103	50 - 150	---	---	1.0	01/25/2021 07:20	01/26/2021 07:30	4819731
LRB	IS-PFOA-13C2	537.1	N/A	--		577770	539839	ng/L	107	50 - 150	---	---	1.0	01/25/2021 07:20	01/26/2021 07:30	4819731
LRB	IS-PFOS-13C4	537.1	N/A	--		276991	271100	ng/L	102	50 - 150	---	---	1.0	01/25/2021 07:20	01/26/2021 07:30	4819731
LRB	SS-NEFOSAA-d5	537.1	N/A	--		135.1720	160	ng/L	84	70 - 130	---	---	1.0	01/25/2021 07:20	01/26/2021 07:30	4819731
LRB	SS-PFDA-13C2	537.1	N/A	--		31.7236	40.0	ng/L	79	70 - 130	---	---	1.0	01/25/2021 07:20	01/26/2021 07:30	4819731
LRB	SS-PFHXA-13C2	537.1	N/A	--		33.2749	40.0	ng/L	83	70 - 130	---	---	1.0	01/25/2021 07:20	01/26/2021 07:30	4819731
LRB	Perfluorobutanesulfonic acid (PFBS)	537.1	2.0	--	<	2.0		ng/L	---	---	---	---	1.0	01/25/2021 07:20	01/26/2021 07:30	4819731
LRB	Perfluorheptanoic acid (PFHpA)	537.1	2.0	--	<	2.0		ng/L	---	---	---	---	1.0	01/25/2021 07:20	01/26/2021 07:30	4819731
LRB	Perfluorohexanesulfonic acid (PFHxS)	537.1	2.0	--	<	2.0		ng/L	---	---	---	---	1.0	01/25/2021 07:20	01/26/2021 07:30	4819731
LRB	Perfluorononanoic acid (PFNA)	537.1	2.0	--	<	2.0		ng/L	---	---	---	---	1.0	01/25/2021 07:20	01/26/2021 07:30	4819731
LRB	Perfluorodecanoic acid (PFDA)	537.1	2.0	--	<	2.0		ng/L	---	---	---	---	1.0	01/25/2021 07:20	01/26/2021 07:30	4819731
LRB	Perfluorohexanoic acid (PFHxA)	537.1	2.0	--	<	2.0		ng/L	---	---	---	---	1.0	01/25/2021 07:20	01/26/2021 07:30	4819731
LRB	Perfluorododecanoic acid (PFDoA)	537.1	2.0	--	<	2.0		ng/L	---	---	---	---	1.0	01/25/2021 07:20	01/26/2021 07:30	4819731

QC Summary Report (cont.)

Sample Type	Analyte	Method	MRL	Client ID	Result Flag	Amount	Target	Units	% Recovery	Recovery Limits	RPD	RPD Limit	Dil Factor	Extracted	Analyzed	EEA ID #
LRB	Perfluorotridecanoic acid (PFTrDA)	537.1	2.0	---	<	2.0		ng/L	---	---	---	---	1.0	01/25/2021 07:20	01/26/2021 07:30	4819731
LRB	Perfluoroundecanoic acid (PFUnA)	537.1	2.0	---	<	2.0		ng/L	---	---	---	---	1.0	01/25/2021 07:20	01/26/2021 07:30	4819731
LRB	N-ethyl Perfluorooctanesulfonamidoacetic acid	537.1	2.0	---	<	2.0		ng/L	---	---	---	---	1.0	01/25/2021 07:20	01/26/2021 07:30	4819731
LRB	N-methyl Perfluorooctanesulfonamidoacetic acid	537.1	2.0	---	<	2.0		ng/L	---	---	---	---	1.0	01/25/2021 07:20	01/26/2021 07:30	4819731
LRB	HFPO-DA/GenX	537.1	2.0	---	<	2.0		ng/L	---	---	---	---	1.0	01/25/2021 07:20	01/26/2021 07:30	4819731
LRB	ADONA	537.1	2.0	---	<	2.0		ng/L	---	---	---	---	1.0	01/25/2021 07:20	01/26/2021 07:30	4819731
LRB	9CI-PF3ONS/F-53B Major	537.1	2.0	---	<	2.0		ng/L	---	---	---	---	1.0	01/25/2021 07:20	01/26/2021 07:30	4819731
LRB	11CI-PF3OUdS/F-53B Minor	537.1	2.0	---	<	2.0		ng/L	---	---	---	---	1.0	01/25/2021 07:20	01/26/2021 07:30	4819731
LRB	Perfluorotetradecanoic acid (PFTeDA)	537.1	2.0	---	<	2.0		ng/L	---	---	---	---	1.0	01/25/2021 07:20	01/26/2021 07:30	4819731
LRB	SS-HFPO-DA-13C3	537.1	N/A	---		33.1490	40.0	ng/L	83	70 - 130	---	---	1.0	01/25/2021 07:20	01/26/2021 07:30	4819731
FBL	Perfluorooctanoic acid (PFOA)	537.1	2.0	---		1.6452	2.0	ng/L	82	50 - 150	---	---	1.0	01/25/2021 07:20	01/26/2021 07:43	4819732
FBL	Perfluorooctanesulfonic acid (PFOS)	537.1	2.0	---		1.6410	2.0	ng/L	82	50 - 150	---	---	1.0	01/25/2021 07:20	01/26/2021 07:43	4819732
FBL	IS-NMeFOSAA-d3	537.1	N/A	---		215229	207534	ng/L	104	50 - 150	---	---	1.0	01/25/2021 07:20	01/26/2021 07:43	4819732
FBL	IS-PFOA-13C2	537.1	N/A	---		588079	539839	ng/L	109	50 - 150	---	---	1.0	01/25/2021 07:20	01/26/2021 07:43	4819732
FBL	IS-PFOS-13C4	537.1	N/A	---		279543	271100	ng/L	103	50 - 150	---	---	1.0	01/25/2021 07:20	01/26/2021 07:43	4819732
FBL	SS-NEFOSAA-d5	537.1	N/A	---		142.6370	160	ng/L	89	70 - 130	---	---	1.0	01/25/2021 07:20	01/26/2021 07:43	4819732
FBL	SS-PFDA-13C2	537.1	N/A	---		35.0508	40.0	ng/L	88	70 - 130	---	---	1.0	01/25/2021 07:20	01/26/2021 07:43	4819732
FBL	SS-PFHxA-13C2	537.1	N/A	---		35.8135	40.0	ng/L	90	70 - 130	---	---	1.0	01/25/2021 07:20	01/26/2021 07:43	4819732
FBL	Perfluorobutanesulfonic acid (PFBS)	537.1	2.0	---		1.2577	2.0	ng/L	63	50 - 150	---	---	1.0	01/25/2021 07:20	01/26/2021 07:43	4819732
FBL	Perfluorohexanoic acid (PFHxA)	537.1	2.0	---		1.4704	2.0	ng/L	74	50 - 150	---	---	1.0	01/25/2021 07:20	01/26/2021 07:43	4819732
FBL	Perfluorohexanesulfonic acid (PFHxS)	537.1	2.0	---		1.4333	2.0	ng/L	72	50 - 150	---	---	1.0	01/25/2021 07:20	01/26/2021 07:43	4819732
FBL	Perfluorononanoic acid (PFNA)	537.1	2.0	---		1.5438	2.0	ng/L	77	50 - 150	---	---	1.0	01/25/2021 07:20	01/26/2021 07:43	4819732
FBL	Perfluorodecanoic acid (PFDA)	537.1	2.0	---		1.3699	2.0	ng/L	68	50 - 150	---	---	1.0	01/25/2021 07:20	01/26/2021 07:43	4819732
FBL	Perfluorohexanoic acid (PFHxA)	537.1	2.0	---		1.3828	2.0	ng/L	69	50 - 150	---	---	1.0	01/25/2021 07:20	01/26/2021 07:43	4819732
FBL	Perfluorododecanoic acid (PFDDA)	537.1	2.0	---		1.3103	2.0	ng/L	66	50 - 150	---	---	1.0	01/25/2021 07:20	01/26/2021 07:43	4819732
FBL	Perfluorotridecanoic acid (PFTrDA)	537.1	2.0	---		1.2350	2.0	ng/L	62	50 - 150	---	---	1.0	01/25/2021 07:20	01/26/2021 07:43	4819732
FBL	Perfluoroundecanoic acid (PFUnA)	537.1	2.0	---		1.3505	2.0	ng/L	68	50 - 150	---	---	1.0	01/25/2021 07:20	01/26/2021 07:43	4819732
FBL	N-ethyl Perfluorooctanesulfonamidoacetic acid	537.1	2.0	---		1.3981	2.0	ng/L	70	50 - 150	---	---	1.0	01/25/2021 07:20	01/26/2021 07:43	4819732
FBL	N-methyl Perfluorooctanesulfonamidoacetic acid	537.1	2.0	---		1.2651	2.0	ng/L	63	50 - 150	---	---	1.0	01/25/2021 07:20	01/26/2021 07:43	4819732
FBL	HFPO-DA/GenX	537.1	2.0	---		1.3482	2.0	ng/L	67	50 - 150	---	---	1.0	01/25/2021 07:20	01/26/2021 07:43	4819732
FBL	ADONA	537.1	2.0	---		1.2846	2.0	ng/L	64	50 - 150	---	---	1.0	01/25/2021 07:20	01/26/2021 07:43	4819732
FBL	9CI-PF3ONS/F-53B Major	537.1	2.0	---		1.5575	2.0	ng/L	78	50 - 150	---	---	1.0	01/25/2021 07:20	01/26/2021 07:43	4819732
FBL	11CI-PF3OUdS/F-53B Minor	537.1	2.0	---		1.4209	2.0	ng/L	71	50 - 150	---	---	1.0	01/25/2021 07:20	01/26/2021 07:43	4819732
FBL	Perfluorotetradecanoic acid (PFTeDA)	537.1	2.0	---		1.2225	2.0	ng/L	61	50 - 150	---	---	1.0	01/25/2021 07:20	01/26/2021 07:43	4819732
FBL	SS-HFPO-DA-13C3	537.1	N/A	---		35.6251	40.0	ng/L	89	70 - 130	---	---	1.0	01/25/2021 07:20	01/26/2021 07:43	4819732
FBM	Perfluorooctanoic acid (PFOA)	537.1	2.0	---		95.3768	100	ng/L	95	70 - 130	---	---	1.0	01/25/2021 07:20	01/26/2021 07:56	4819735
FBM	Perfluorooctanesulfonic acid (PFOS)	537.1	2.0	---		93.8212	100	ng/L	94	70 - 130	---	---	1.0	01/25/2021 07:20	01/26/2021 07:56	4819735
FBM	IS-NMeFOSAA-d3	537.1	N/A	---		202153	207534	ng/L	97	50 - 150	---	---	1.0	01/25/2021 07:20	01/26/2021 07:56	4819735
FBM	IS-PFOA-13C2	537.1	N/A	---		555479	539839	ng/L	103	50 - 150	---	---	1.0	01/25/2021 07:20	01/26/2021 07:56	4819735
FBM	IS-PFOS-13C4	537.1	N/A	---		258874	271100	ng/L	95	50 - 150	---	---	1.0	01/25/2021 07:20	01/26/2021 07:56	4819735
FBM	SS-NEFOSAA-d5	537.1	N/A	---		146.7770	160	ng/L	92	70 - 130	---	---	1.0	01/25/2021 07:20	01/26/2021 07:56	4819735

QC Summary Report (cont.)

Sample Type	Analyte	Method	MRL	Client ID	Result Flag	Amount	Target	Units	% Recovery	Recovery Limits	RPD	RPD Limit	Dil Factor	Extracted	Analyzed	EEA ID #
FBM	SS-PFDA-13C2	537.1	N/A	---		35.9671	40.0	ng/L	90	70 - 130	---	---	1.0	01/25/2021 07:20	01/26/2021 07:56	4819735
FBM	SS-PFHXA-13C2	537.1	N/A	---		34.7729	40.0	ng/L	87	70 - 130	---	---	1.0	01/25/2021 07:20	01/26/2021 07:56	4819735
FBM	Perfluorobutanesulfonic acid (PFBS)	537.1	2.0	---		78.6721	100	ng/L	79	70 - 130	---	---	1.0	01/25/2021 07:20	01/26/2021 07:56	4819735
FBM	Perfluoroheptanoic acid (PFHpA)	537.1	2.0	---		91.1963	100	ng/L	91	70 - 130	---	---	1.0	01/25/2021 07:20	01/26/2021 07:56	4819735
FBM	Perfluorohexanesulfonic acid (PFHxS)	537.1	2.0	---		96.9995	100	ng/L	97	70 - 130	---	---	1.0	01/25/2021 07:20	01/26/2021 07:56	4819735
FBM	Perfluorononanoic acid (PFNA)	537.1	2.0	---		91.1316	100	ng/L	91	70 - 130	---	---	1.0	01/25/2021 07:20	01/26/2021 07:56	4819735
FBM	Perfluorodecanoic acid (PFDA)	537.1	2.0	---		82.9744	100	ng/L	83	70 - 130	---	---	1.0	01/25/2021 07:20	01/26/2021 07:56	4819735
FBM	Perfluorohexanoic acid (PFHxA)	537.1	2.0	---		82.2832	100	ng/L	82	70 - 130	---	---	1.0	01/25/2021 07:20	01/26/2021 07:56	4819735
FBM	Perfluorododecanoic acid (PFDoA)	537.1	2.0	---		80.4903	100	ng/L	80	70 - 130	---	---	1.0	01/25/2021 07:20	01/26/2021 07:56	4819735
FBM	Perfluorotridecanoic acid (PFTtDA)	537.1	2.0	---		77.5354	100	ng/L	78	70 - 130	---	---	1.0	01/25/2021 07:20	01/26/2021 07:56	4819735
FBM	Perfluoroundecanoic acid (PFUnA)	537.1	2.0	---		82.0935	100	ng/L	82	70 - 130	---	---	1.0	01/25/2021 07:20	01/26/2021 07:56	4819735
FBM	N-ethyl Perfluorooctanesulfonamidoacetic acid	537.1	2.0	---		87.1956	100	ng/L	87	70 - 130	---	---	1.0	01/25/2021 07:20	01/26/2021 07:56	4819735
FBM	N-methyl Perfluorooctanesulfonamidoacetic acid	537.1	2.0	---		89.2757	100	ng/L	89	70 - 130	---	---	1.0	01/25/2021 07:20	01/26/2021 07:56	4819735
FBM	HFPO-DA/GenX	537.1	2.0	---		81.0130	100	ng/L	81	70 - 130	---	---	1.0	01/25/2021 07:20	01/26/2021 07:56	4819735
FBM	ADONA	537.1	2.0	---		85.9834	100	ng/L	86	70 - 130	---	---	1.0	01/25/2021 07:20	01/26/2021 07:56	4819735
FBM	9C+PF3ONS/F-53B Major	537.1	2.0	---		94.3860	100	ng/L	94	70 - 130	---	---	1.0	01/25/2021 07:20	01/26/2021 07:56	4819735
FBM	11C+PF3OudS/F-53B Minor	537.1	2.0	---		88.2583	100	ng/L	88	70 - 130	---	---	1.0	01/25/2021 07:20	01/26/2021 07:56	4819735
FBM	Perfluorotetradecanoic acid (PFTtDA)	537.1	2.0	---		76.3615	100	ng/L	76	70 - 130	---	---	1.0	01/25/2021 07:20	01/26/2021 07:56	4819735
FBM	SS-HFPO-DA-13C3	537.1	N/A	---		34.8395	40.0	ng/L	87	70 - 130	---	---	1.0	01/25/2021 07:20	01/26/2021 07:56	4819735
CCM	Perfluorooctanoic acid (PFOA)	537.1	2.0	---		99.1221	100	ng/L	99	70 - 130	---	---	1.0	01/22/2021 08:59	01/26/2021 11:17	4819837
CCM	Perfluorooctanesulfonic acid (PFOS)	537.1	2.0	---		99.7256	100	ng/L	100	70 - 130	---	---	1.0	01/22/2021 08:59	01/26/2021 11:17	4819837
CCM	IS-NMeFOSAA-d3	537.1	N/A	---		156959	156959	ng/L	100	50 - 150	---	---	1.0	01/22/2021 08:59	01/26/2021 11:17	4819837
CCM	IS-PFOA-13C2	537.1	N/A	---		421169	421169	ng/L	100	50 - 150	---	---	1.0	01/22/2021 08:59	01/26/2021 11:17	4819837
CCM	IS-PFOS-13C4	537.1	N/A	---		208065	208065	ng/L	100	50 - 150	---	---	1.0	01/22/2021 08:59	01/26/2021 11:17	4819837
CCM	SS-NEFOSAA-d5	537.1	N/A	---		163.6040	160	ng/L	102	70 - 130	---	---	1.0	01/22/2021 08:59	01/26/2021 11:17	4819837
CCM	SS-PFDA-13C2	537.1	N/A	---		39.9130	40.0	ng/L	100	70 - 130	---	---	1.0	01/22/2021 08:59	01/26/2021 11:17	4819837
CCM	SS-PFHXA-13C2	537.1	N/A	---		47.1902	40.0	ng/L	118	70 - 130	---	---	1.0	01/22/2021 08:59	01/26/2021 11:17	4819837
CCM	Perfluorobutanesulfonic acid (PFBS)	537.1	2.0	---		108.4640	100	ng/L	108	70 - 130	---	---	1.0	01/22/2021 08:59	01/26/2021 11:17	4819837
CCM	Perfluoroheptanoic acid (PFHpA)	537.1	2.0	---		105.8230	100	ng/L	106	70 - 130	---	---	1.0	01/22/2021 08:59	01/26/2021 11:17	4819837
CCM	Perfluorohexanesulfonic acid (PFHxS)	537.1	2.0	---		107.1550	100	ng/L	107	70 - 130	---	---	1.0	01/22/2021 08:59	01/26/2021 11:17	4819837
CCM	Perfluorononanoic acid (PFNA)	537.1	2.0	---		98.5238	100	ng/L	99	70 - 130	---	---	1.0	01/22/2021 08:59	01/26/2021 11:17	4819837
CCM	Perfluorodecanoic acid (PFDA)	537.1	2.0	---		95.3363	100	ng/L	95	70 - 130	---	---	1.0	01/22/2021 08:59	01/26/2021 11:17	4819837
CCM	Perfluorohexanoic acid (PFHxA)	537.1	2.0	---		118.1250	100	ng/L	118	70 - 130	---	---	1.0	01/22/2021 08:59	01/26/2021 11:17	4819837
CCM	Perfluorododecanoic acid (PFDoA)	537.1	2.0	---		98.5698	100	ng/L	99	70 - 130	---	---	1.0	01/22/2021 08:59	01/26/2021 11:17	4819837
CCM	Perfluorotridecanoic acid (PFTtDA)	537.1	2.0	---		103.3240	100	ng/L	103	70 - 130	---	---	1.0	01/22/2021 08:59	01/26/2021 11:17	4819837
CCM	Perfluoroundecanoic acid (PFUnA)	537.1	2.0	---		97.0904	100	ng/L	97	70 - 130	---	---	1.0	01/22/2021 08:59	01/26/2021 11:17	4819837
CCM	N-ethyl Perfluorooctanesulfonamidoacetic acid	537.1	2.0	---		103.5060	100	ng/L	104	70 - 130	---	---	1.0	01/22/2021 08:59	01/26/2021 11:17	4819837
CCM	N-methyl Perfluorooctanesulfonamidoacetic acid	537.1	2.0	---		101.2350	100	ng/L	101	70 - 130	---	---	1.0	01/22/2021 08:59	01/26/2021 11:17	4819837
CCM	HFPO-DA/GenX	537.1	2.0	---		119.2500	100	ng/L	119	70 - 130	---	---	1.0	01/22/2021 08:59	01/26/2021 11:17	4819837
CCM	ADONA	537.1	2.0	---		101.6050	100	ng/L	102	70 - 130	---	---	1.0	01/22/2021 08:59	01/26/2021 11:17	4819837
CCM	9C+PF3ONS/F-53B Major	537.1	2.0	---		98.4794	100	ng/L	98	70 - 130	---	---	1.0	01/22/2021 08:59	01/26/2021 11:17	4819837

QC Summary Report (cont.)

Sample Type	Analyte	Method	MRL	Client ID	Result Flag	Amount	Target	Units	% Recovery	Recovery Limits	RPD	RPD Limit	Dil Factor	Extracted	Analyzed	EEA ID #
CCM	11Cl-PF3OUdF-53B Minor	537.1	2.0	---		102.3970	100	ng/L	102	70 - 130	---	---	1.0	01/22/2021 08:59	01/26/2021 11:17	4819837
CCM	Perfluorotetradecanoic acid (PFTeDA)	537.1	2.0	---		108.6670	100	ng/L	109	70 - 130	---	---	1.0	01/22/2021 08:59	01/26/2021 11:17	4819837
CCM	SS-HFO-DA-13C3	537.1	N/A	---		50.8945	40.0	ng/L	127	70 - 130	---	---	1.0	01/22/2021 08:59	01/26/2021 11:17	4819837
CCH	Perfluorooctanoic acid (PFOA)	537.1	2.0	---		197.3080	200	ng/L	99	70 - 130	---	---	1.0	01/22/2021 08:59	01/26/2021 15:01	4819841
CCH	Perfluorooctanesulfonic acid (PFOS)	537.1	2.0	---		201.7700	200	ng/L	101	70 - 130	---	---	1.0	01/22/2021 08:59	01/26/2021 15:01	4819841
CCH	IS-NMeFOSAA-d3	537.1	N/A	---		201211	201211	ng/L	100	50 - 150	---	---	1.0	01/22/2021 08:59	01/26/2021 15:01	4819841
CCH	IS-PFOA-13C2	537.1	N/A	---		485655	485655	ng/L	100	50 - 150	---	---	1.0	01/22/2021 08:59	01/26/2021 15:01	4819841
CCH	IS-PFOS-13C4	537.1	N/A	---		243247	243247	ng/L	100	50 - 150	---	---	1.0	01/22/2021 08:59	01/26/2021 15:01	4819841
CCH	SS-NEFOSAA-d5	537.1	N/A	---		156.0550	160	ng/L	98	70 - 130	---	---	1.0	01/22/2021 08:59	01/26/2021 15:01	4819841
CCH	SS-PFDA-13C2	537.1	N/A	---		40.4108	40.0	ng/L	101	70 - 130	---	---	1.0	01/22/2021 08:59	01/26/2021 15:01	4819841
CCH	SS-PFHXA-13C2	537.1	N/A	---		40.9906	40.0	ng/L	102	70 - 130	---	---	1.0	01/22/2021 08:59	01/26/2021 15:01	4819841
CCH	Perfluorobutanesulfonic acid (PFBS)	537.1	2.0	---		200.6690	200	ng/L	100	70 - 130	---	---	1.0	01/22/2021 08:59	01/26/2021 15:01	4819841
CCH	Perfluoroheptanoic acid (PFHpA)	537.1	2.0	---		199.7430	200	ng/L	100	70 - 130	---	---	1.0	01/22/2021 08:59	01/26/2021 15:01	4819841
CCH	Perfluorohexanesulfonic acid (PFHxS)	537.1	2.0	---		203.9720	200	ng/L	102	70 - 130	---	---	1.0	01/22/2021 08:59	01/26/2021 15:01	4819841
CCH	Perfluorononanoic acid (PFNA)	537.1	2.0	---		196.8200	200	ng/L	98	70 - 130	---	---	1.0	01/22/2021 08:59	01/26/2021 15:01	4819841
CCH	Perfluorodecanoic acid (PFDA)	537.1	2.0	---		192.9720	200	ng/L	96	70 - 130	---	---	1.0	01/22/2021 08:59	01/26/2021 15:01	4819841
CCH	Perfluorohexanoic acid (PFHxA)	537.1	2.0	---		203.0480	200	ng/L	102	70 - 130	---	---	1.0	01/22/2021 08:59	01/26/2021 15:01	4819841
CCH	Perfluorododecanoic acid (PFDoA)	537.1	2.0	---		191.4650	200	ng/L	96	70 - 130	---	---	1.0	01/22/2021 08:59	01/26/2021 15:01	4819841
CCH	Perfluorotridecanoic acid (PFTDA)	537.1	2.0	---		199.5450	200	ng/L	100	70 - 130	---	---	1.0	01/22/2021 08:59	01/26/2021 15:01	4819841
CCH	Perfluoroundecanoic acid (PFUnA)	537.1	2.0	---		197.8540	200	ng/L	99	70 - 130	---	---	1.0	01/22/2021 08:59	01/26/2021 15:01	4819841
CCH	N-ethyl Perfluorooctanesulfonamidoacetic acid	537.1	2.0	---		200.0010	200	ng/L	100	70 - 130	---	---	1.0	01/22/2021 08:59	01/26/2021 15:01	4819841
CCH	N-methyl Perfluorooctanesulfonamidoacetic acid	537.1	2.0	---		200.5600	200	ng/L	100	70 - 130	---	---	1.0	01/22/2021 08:59	01/26/2021 15:01	4819841
CCH	HFO-DA/GenX	537.1	2.0	---		197.8630	200	ng/L	99	70 - 130	---	---	1.0	01/22/2021 08:59	01/26/2021 15:01	4819841
CCH	ADONA	537.1	2.0	---		203.2440	200	ng/L	102	70 - 130	---	---	1.0	01/22/2021 08:59	01/26/2021 15:01	4819841
CCH	9Cl-PF3ONSIF-53B Major	537.1	2.0	---		199.5170	200	ng/L	100	70 - 130	---	---	1.0	01/22/2021 08:59	01/26/2021 15:01	4819841
CCH	11Cl-PF3OUdF-53B Minor	537.1	2.0	---		194.6670	200	ng/L	97	70 - 130	---	---	1.0	01/22/2021 08:59	01/26/2021 15:01	4819841
CCH	Perfluorotetradecanoic acid (PFTeDA)	537.1	2.0	---		260.3010	200	ng/L	130	70 - 130	---	---	1.0	01/22/2021 08:59	01/26/2021 15:01	4819841
CCH	SS-HFO-DA-13C3	537.1	N/A	---		41.6684	40.0	ng/L	104	70 - 130	---	---	1.0	01/22/2021 08:59	01/26/2021 15:01	4819841
CCM	Perfluorooctanoic acid (PFOA)	537.1	2.0	---		100.1220	100	ng/L	100	70 - 130	---	---	1.0	01/22/2021 08:59	01/26/2021 17:24	4820887
CCM	Perfluorooctanesulfonic acid (PFOS)	537.1	2.0	---		100.7410	100	ng/L	101	70 - 130	---	---	1.0	01/22/2021 08:59	01/26/2021 17:24	4820887
CCM	IS-NMeFOSAA-d3	537.1	N/A	---		214656	214656	ng/L	100	50 - 150	---	---	1.0	01/22/2021 08:59	01/26/2021 17:24	4820887
CCM	IS-PFOA-13C2	537.1	N/A	---		551837	551837	ng/L	100	50 - 150	---	---	1.0	01/22/2021 08:59	01/26/2021 17:24	4820887
CCM	IS-PFOS-13C4	537.1	N/A	---		272427	272427	ng/L	100	50 - 150	---	---	1.0	01/22/2021 08:59	01/26/2021 17:24	4820887
CCM	SS-NEFOSAA-d5	537.1	N/A	---		162.0090	160	ng/L	101	70 - 130	---	---	1.0	01/22/2021 08:59	01/26/2021 17:24	4820887
CCM	SS-PFDA-13C2	537.1	N/A	---		39.6925	40.0	ng/L	99	70 - 130	---	---	1.0	01/22/2021 08:59	01/26/2021 17:24	4820887
CCM	SS-PFHXA-13C2	537.1	N/A	---		38.1841	40.0	ng/L	95	70 - 130	---	---	1.0	01/22/2021 08:59	01/26/2021 17:24	4820887
CCM	Perfluorobutanesulfonic acid (PFBS)	537.1	2.0	---		98.2408	100	ng/L	98	70 - 130	---	---	1.0	01/22/2021 08:59	01/26/2021 17:24	4820887
CCM	Perfluoroheptanoic acid (PFHpA)	537.1	2.0	---		98.6621	100	ng/L	99	70 - 130	---	---	1.0	01/22/2021 08:59	01/26/2021 17:24	4820887
CCM	Perfluorohexanesulfonic acid (PFHxS)	537.1	2.0	---		101.0780	100	ng/L	101	70 - 130	---	---	1.0	01/22/2021 08:59	01/26/2021 17:24	4820887
CCM	Perfluorononanoic acid (PFNA)	537.1	2.0	---		101.8220	100	ng/L	102	70 - 130	---	---	1.0	01/22/2021 08:59	01/26/2021 17:24	4820887
CCM	Perfluorodecanoic acid (PFDA)	537.1	2.0	---		99.0124	100	ng/L	99	70 - 130	---	---	1.0	01/22/2021 08:59	01/26/2021 17:24	4820887

QC Summary Report (cont.)

Sample Type	Analyte	Method	MRL	Client ID	Result Flag	Amount	Target	Units	% Recovery	Recovery Limits	RPD	RPD Limit	Dil Factor	Extracted	Analyzed	EEA ID #
CCM	Perfluorohexanoic acid (PFHxA)	537.1	2.0	---		95.5166	100	ng/L	96	70 - 130	---	---	1.0	01/22/2021 08:59	01/26/2021 17:24	4820887
CCM	Perfluorododecanoic acid (PFDoA)	537.1	2.0	---		99.6902	100	ng/L	100	70 - 130	---	---	1.0	01/22/2021 08:59	01/26/2021 17:24	4820887
CCM	Perfluorotridecanoic acid (PFTriDA)	537.1	2.0	---		96.3583	100	ng/L	96	70 - 130	---	---	1.0	01/22/2021 08:59	01/26/2021 17:24	4820887
CCM	Perfluoroundecanoic acid (PFUnA)	537.1	2.0	---		99.4118	100	ng/L	99	70 - 130	---	---	1.0	01/22/2021 08:59	01/26/2021 17:24	4820887
CCM	N-ethyl Perfluorooctanesulfonamidoacetic acid	537.1	2.0	---		103.7050	100	ng/L	104	70 - 130	---	---	1.0	01/22/2021 08:59	01/26/2021 17:24	4820887
CCM	N-methyl Perfluorooctanesulfonamidoacetic acid	537.1	2.0	---		100.6650	100	ng/L	101	70 - 130	---	---	1.0	01/22/2021 08:59	01/26/2021 17:24	4820887
CCM	HFPO-DA/GenX	537.1	2.0	---		94.5764	100	ng/L	95	70 - 130	---	---	1.0	01/22/2021 08:59	01/26/2021 17:24	4820887
CCM	ADONA	537.1	2.0	---		95.4550	100	ng/L	95	70 - 130	---	---	1.0	01/22/2021 08:59	01/26/2021 17:24	4820887
CCM	9Cl-PF3ONS/F-53B Major	537.1	2.0	---		104.5260	100	ng/L	105	70 - 130	---	---	1.0	01/22/2021 08:59	01/26/2021 17:24	4820887
CCM	11Cl-PF3OUdS/F-53B Minor	537.1	2.0	---		103.4770	100	ng/L	103	70 - 130	---	---	1.0	01/22/2021 08:59	01/26/2021 17:24	4820887
CCM	Perfluorotetradecanoic acid (PFTeDA)	537.1	2.0	---		97.1024	100	ng/L	97	70 - 130	---	---	1.0	01/22/2021 08:59	01/26/2021 17:24	4820887
CCM	SS-HFO-DA-13C3	537.1	N/A	---		38.1817	40.0	ng/L	95	70 - 130	---	---	1.0	01/22/2021 08:59	01/26/2021 17:24	4820887
CCH	Perfluorooctanoic acid (PFOA)	537.1	2.0	---		197.4910	200	ng/L	99	70 - 130	---	---	1.0	01/22/2021 08:59	01/27/2021 09:21	4820890
CCH	Perfluorooctanesulfonic acid (PFOS)	537.1	2.0	---		197.6440	200	ng/L	99	70 - 130	---	---	1.0	01/22/2021 08:59	01/27/2021 09:21	4820890
CCH	IS-NMeFOSAA-d3	537.1	N/A	---		199589	199589	ng/L	100	50 - 150	---	---	1.0	01/22/2021 08:59	01/27/2021 09:21	4820890
CCH	IS-PFOA-13C2	537.1	N/A	---		475914	475914	ng/L	100	50 - 150	---	---	1.0	01/22/2021 08:59	01/27/2021 09:21	4820890
CCH	IS-PFOS-13C4	537.1	N/A	---		249909	249909	ng/L	100	50 - 150	---	---	1.0	01/22/2021 08:59	01/27/2021 09:21	4820890
CCH	SS-NEFOSAA-d5	537.1	N/A	---		154.8720	160	ng/L	97	70 - 130	---	---	1.0	01/22/2021 08:59	01/27/2021 09:21	4820890
CCH	SS-PFDA-13C2	537.1	N/A	---		41.5787	40.0	ng/L	104	70 - 130	---	---	1.0	01/22/2021 08:59	01/27/2021 09:21	4820890
CCH	SS-PFHxA-13C2	537.1	N/A	---		42.6928	40.0	ng/L	107	70 - 130	---	---	1.0	01/22/2021 08:59	01/27/2021 09:21	4820890
CCH	Perfluorobutanesulfonic acid (PFBS)	537.1	2.0	---		198.7910	200	ng/L	99	70 - 130	---	---	1.0	01/22/2021 08:59	01/27/2021 09:21	4820890
CCH	Perfluorohexanoic acid (PFHxA)	537.1	2.0	---		202.1920	200	ng/L	101	70 - 130	---	---	1.0	01/22/2021 08:59	01/27/2021 09:21	4820890
CCH	Perfluorohexanesulfonic acid (PFHxS)	537.1	2.0	---		195.3460	200	ng/L	98	70 - 130	---	---	1.0	01/22/2021 08:59	01/27/2021 09:21	4820890
CCH	Perfluorononanoic acid (PFNA)	537.1	2.0	---		203.7410	200	ng/L	102	70 - 130	---	---	1.0	01/22/2021 08:59	01/27/2021 09:21	4820890
CCH	Perfluorodecanoic acid (PFDA)	537.1	2.0	---		198.5020	200	ng/L	99	70 - 130	---	---	1.0	01/22/2021 08:59	01/27/2021 09:21	4820890
CCH	Perfluorohexanoic acid (PFHxA)	537.1	2.0	---		210.9930	200	ng/L	105	70 - 130	---	---	1.0	01/22/2021 08:59	01/27/2021 09:21	4820890
CCH	Perfluorododecanoic acid (PFDoA)	537.1	2.0	---		200.7550	200	ng/L	100	70 - 130	---	---	1.0	01/22/2021 08:59	01/27/2021 09:21	4820890
CCH	Perfluorotridecanoic acid (PFTriDA)	537.1	2.0	---		201.2530	200	ng/L	101	70 - 130	---	---	1.0	01/22/2021 08:59	01/27/2021 09:21	4820890
CCH	Perfluoroundecanoic acid (PFUnA)	537.1	2.0	---		207.7650	200	ng/L	104	70 - 130	---	---	1.0	01/22/2021 08:59	01/27/2021 09:21	4820890
CCH	N-ethyl Perfluorooctanesulfonamidoacetic acid	537.1	2.0	---		197.7470	200	ng/L	99	70 - 130	---	---	1.0	01/22/2021 08:59	01/27/2021 09:21	4820890
CCH	N-methyl Perfluorooctanesulfonamidoacetic acid	537.1	2.0	---		199.6920	200	ng/L	100	70 - 130	---	---	1.0	01/22/2021 08:59	01/27/2021 09:21	4820890
CCH	HFPO-DA/GenX	537.1	2.0	---		214.0380	200	ng/L	107	70 - 130	---	---	1.0	01/22/2021 08:59	01/27/2021 09:21	4820890
CCH	ADONA	537.1	2.0	---		208.7450	200	ng/L	104	70 - 130	---	---	1.0	01/22/2021 08:59	01/27/2021 09:21	4820890
CCH	9Cl-PF3ONS/F-53B Major	537.1	2.0	---		196.9720	200	ng/L	98	70 - 130	---	---	1.0	01/22/2021 08:59	01/27/2021 09:21	4820890
CCH	11Cl-PF3OUdS/F-53B Minor	537.1	2.0	---		190.6050	200	ng/L	95	70 - 130	---	---	1.0	01/22/2021 08:59	01/27/2021 09:21	4820890
CCH	Perfluorotetradecanoic acid (PFTeDA)	537.1	2.0	---		192.5870	200	ng/L	96	70 - 130	---	---	1.0	01/22/2021 08:59	01/27/2021 09:21	4820890
CCH	SS-HFO-DA-13C3	537.1	N/A	---		43.9733	40.0	ng/L	110	70 - 130	---	---	1.0	01/22/2021 08:59	01/27/2021 09:21	4820890
CCM	Perfluorooctanoic acid (PFOA)	537.1	2.0	---		98.5583	100	ng/L	99	70 - 130	---	---	1.0	01/22/2021 08:59	01/27/2021 09:49	4820888
CCM	Perfluorooctanesulfonic acid (PFOS)	537.1	2.0	---		99.9623	100	ng/L	100	70 - 130	---	---	1.0	01/22/2021 08:59	01/27/2021 09:49	4820888
CCM	IS-NMeFOSAA-d3	537.1	N/A	---		191470	191470	ng/L	100	50 - 150	---	---	1.0	01/22/2021 08:59	01/27/2021 09:49	4820888
CCM	IS-PFOA-13C2	537.1	N/A	---		474095	474095	ng/L	100	50 - 150	---	---	1.0	01/22/2021 08:59	01/27/2021 09:49	4820888

QC Summary Report (cont.)

Sample Type	Analyte	Method	MRL	Client ID	Result Flag	Amount	Target	Units	% Recovery	Recovery Limits	RPD	RPD Limit	Dil Factor	Extracted	Analyzed	EEA ID #
CCM	IS-PFOS-13C4	537.1	N/A	---		244252	244252	ng/L	100	50 - 150	---	---	1.0	01/22/2021 08:59	01/27/2021 09:49	4820888
CCM	SS-NEFOSAA-d5	537.1	N/A	---		160.5770	160	ng/L	100	70 - 130	---	---	1.0	01/22/2021 08:59	01/27/2021 09:49	4820888
CCM	SS-PFDA-13C2	537.1	N/A	---		40.7170	40.0	ng/L	102	70 - 130	---	---	1.0	01/22/2021 08:59	01/27/2021 09:49	4820888
CCM	SS-PFHXA-13C2	537.1	N/A	---		43.1226	40.0	ng/L	108	70 - 130	---	---	1.0	01/22/2021 08:59	01/27/2021 09:49	4820888
CCM	Perfluorobutanesulfonic acid (PFBS)	537.1	2.0	---		99.0189	100	ng/L	99	70 - 130	---	---	1.0	01/22/2021 08:59	01/27/2021 09:49	4820888
CCM	Perfluoroheptanoic acid (PFHpA)	537.1	2.0	---		99.7800	100	ng/L	100	70 - 130	---	---	1.0	01/22/2021 08:59	01/27/2021 09:49	4820888
CCM	Perfluorohexanesulfonic acid (PFHxS)	537.1	2.0	---		99.6002	100	ng/L	100	70 - 130	---	---	1.0	01/22/2021 08:59	01/27/2021 09:49	4820888
CCM	Perfluorononanoic acid (PFNA)	537.1	2.0	---		100.8770	100	ng/L	101	70 - 130	---	---	1.0	01/22/2021 08:59	01/27/2021 09:49	4820888
CCM	Perfluorodecanoic acid (PFDA)	537.1	2.0	---		101.4390	100	ng/L	101	70 - 130	---	---	1.0	01/22/2021 08:59	01/27/2021 09:49	4820888
CCM	Perfluorohexanoic acid (PFHxA)	537.1	2.0	---		107.5380	100	ng/L	108	70 - 130	---	---	1.0	01/22/2021 08:59	01/27/2021 09:49	4820888
CCM	Perfluorododecanoic acid (PFDoA)	537.1	2.0	---		102.2880	100	ng/L	102	70 - 130	---	---	1.0	01/22/2021 08:59	01/27/2021 09:49	4820888
CCM	Perfluorotridecanoic acid (PFTtDA)	537.1	2.0	---		102.5890	100	ng/L	103	70 - 130	---	---	1.0	01/22/2021 08:59	01/27/2021 09:49	4820888
CCM	Perfluoroundecanoic acid (PFUnA)	537.1	2.0	---		103.3040	100	ng/L	103	70 - 130	---	---	1.0	01/22/2021 08:59	01/27/2021 09:49	4820888
CCM	N-ethyl Perfluorooctanesulfonamidoacetic acid	537.1	2.0	---		102.6260	100	ng/L	103	70 - 130	---	---	1.0	01/22/2021 08:59	01/27/2021 09:49	4820888
CCM	N-methyl Perfluorooctanesulfonamidoacetic acid	537.1	2.0	---		100.1300	100	ng/L	100	70 - 130	---	---	1.0	01/22/2021 08:59	01/27/2021 09:49	4820888
CCM	HFPO-DA/GenX	537.1	2.0	---		104.2720	100	ng/L	104	70 - 130	---	---	1.0	01/22/2021 08:59	01/27/2021 09:49	4820888
CCM	ADONA	537.1	2.0	---		97.8180	100	ng/L	98	70 - 130	---	---	1.0	01/22/2021 08:59	01/27/2021 09:49	4820888
CCM	9C-PF3ONS/F-53B Major	537.1	2.0	---		101.5120	100	ng/L	102	70 - 130	---	---	1.0	01/22/2021 08:59	01/27/2021 09:49	4820888
CCM	11C-PF3OUdS/F-53B Minor	537.1	2.0	---		101.9620	100	ng/L	102	70 - 130	---	---	1.0	01/22/2021 08:59	01/27/2021 09:49	4820888
CCM	Perfluorotetradecanoic acid (PFTeDA)	537.1	2.0	---		112.4320	100	ng/L	112	70 - 130	---	---	1.0	01/22/2021 08:59	01/27/2021 09:49	4820888
CCM	SS-HFPO-DA-13C3	537.1	N/A	---		44.4808	40.0	ng/L	111	70 - 130	---	---	1.0	01/22/2021 08:59	01/27/2021 09:49	4820888

Eurofins Eaton Analytical

Run Log

Run ID: 284766 Method: 537.1

<u>Type</u>	<u>Sample Id</u>	<u>Sample Site</u>	<u>Matrix</u>	<u>Instrument ID</u>	<u>Analysis Date</u>	<u>Calibration File</u>
CCL	4820250		OS	FL	01/26/2021 19:08	012621M537_1b-FL.mdb
LRB	4820237		RW	FL	01/26/2021 19:34	012621M537_1b-FL.mdb
FBL	4820238		RW	FL	01/26/2021 19:47	012621M537_1b-FL.mdb
FBH	4820241		RW	FL	01/26/2021 20:00	012621M537_1b-FL.mdb
CCM	4820251		OS	FL	01/27/2021 00:21	012621M537_1b-FL.mdb
CCH	4820252		OS	FL	01/27/2021 02:44	012621M537_1b-FL.mdb
CCM	4820921		OS	FL	01/27/2021 10:08	012621M537_1b-FL.mdb
CCH	4820920		OS	FL	01/27/2021 11:20	012621M537_1b-FL.mdb

QC Summary Report

Sample Type	Analyte	Method	MDA95	Client ID	Result Flag	Amount	Target	Units	% Recovery	Recovery Limits	RPD	RPD Limit	Dil Factor	Extracted	Analyzed	EEA ID #
CCL	Perfluorooctanoic acid (PFOA)	537.1	2.0	--		1.7404	2.0	ng/L	87	50 - 150	---	---	1.0	01/26/2021 08:41	01/26/2021 19:08	4820250
CCL	Perfluorooctanesulfonic acid (PFOS)	537.1	2.0	--		1.8185	2.0	ng/L	91	50 - 150	---	---	1.0	01/26/2021 08:41	01/26/2021 19:08	4820250
CCL	IS-NMeFOSAA-d3	537.1	N/A	--		174812	174812	ng/L	100	50 - 150	---	---	1.0	01/26/2021 08:41	01/26/2021 19:08	4820250
CCL	IS-PFOA-13C2	537.1	N/A	--		449813	449813	ng/L	100	50 - 150	---	---	1.0	01/26/2021 08:41	01/26/2021 19:08	4820250
CCL	IS-PFOS-13C4	537.1	N/A	--		227165	227165	ng/L	100	50 - 150	---	---	1.0	01/26/2021 08:41	01/26/2021 19:08	4820250
CCL	SS-NEFOSAA-d5	537.1	N/A	--		162.3760	160	ng/L	101	70 - 130	---	---	1.0	01/26/2021 08:41	01/26/2021 19:08	4820250
CCL	SS-PFDA-13C2	537.1	N/A	--		38.8989	40.0	ng/L	97	70 - 130	---	---	1.0	01/26/2021 08:41	01/26/2021 19:08	4820250
CCL	SS-PFHXA-13C2	537.1	N/A	--		41.7570	40.0	ng/L	104	70 - 130	---	---	1.0	01/26/2021 08:41	01/26/2021 19:08	4820250
CCL	Perfluorobutanesulfonic acid (PFBS)	537.1	2.0	--		1.5318	2.0	ng/L	77	50 - 150	---	---	1.0	01/26/2021 08:41	01/26/2021 19:08	4820250
CCL	Perfluorheptanoic acid (PFHpA)	537.1	2.0	--		1.5996	2.0	ng/L	80	50 - 150	---	---	1.0	01/26/2021 08:41	01/26/2021 19:08	4820250
CCL	Perfluorohexanesulfonic acid (PFHxS)	537.1	2.0	--		1.5316	2.0	ng/L	77	50 - 150	---	---	1.0	01/26/2021 08:41	01/26/2021 19:08	4820250
CCL	Perfluorononanoic acid (PFNA)	537.1	2.0	--		1.7418	2.0	ng/L	87	50 - 150	---	---	1.0	01/26/2021 08:41	01/26/2021 19:08	4820250
CCL	Perfluorodecanoic acid (PFDA)	537.1	2.0	--		1.7815	2.0	ng/L	89	50 - 150	---	---	1.0	01/26/2021 08:41	01/26/2021 19:08	4820250
CCL	Perfluorohexanoic acid (PFHxA)	537.1	2.0	--		1.7425	2.0	ng/L	87	50 - 150	---	---	1.0	01/26/2021 08:41	01/26/2021 19:08	4820250
CCL	Perfluorododecanoic acid (PFDoA)	537.1	2.0	--		1.9384	2.0	ng/L	97	50 - 150	---	---	1.0	01/26/2021 08:41	01/26/2021 19:08	4820250
CCL	Perfluorotridecanoic acid (PFTDA)	537.1	2.0	--		1.9521	2.0	ng/L	98	50 - 150	---	---	1.0	01/26/2021 08:41	01/26/2021 19:08	4820250
CCL	Perfluoroundecanoic acid (PFUnA)	537.1	2.0	--		1.8209	2.0	ng/L	91	50 - 150	---	---	1.0	01/26/2021 08:41	01/26/2021 19:08	4820250
CCL	N-ethyl Perfluorooctanesulfonamidoacetic acid	537.1	2.0	--		1.8873	2.0	ng/L	94	50 - 150	---	---	1.0	01/26/2021 08:41	01/26/2021 19:08	4820250
CCL	N-methyl Perfluorooctanesulfonamidoacetic acid	537.1	2.0	--		1.7863	2.0	ng/L	89	50 - 150	---	---	1.0	01/26/2021 08:41	01/26/2021 19:08	4820250
CCL	HFPO-DA/GenX	537.1	2.0	--		1.6173	2.0	ng/L	81	50 - 150	---	---	1.0	01/26/2021 08:41	01/26/2021 19:08	4820250
CCL	ADONA	537.1	2.0	--		1.4027	2.0	ng/L	70	50 - 150	---	---	1.0	01/26/2021 08:41	01/26/2021 19:08	4820250
CCL	9CI-PF3ONS/F-53B Major	537.1	2.0	--		1.9888	2.0	ng/L	99	50 - 150	---	---	1.0	01/26/2021 08:41	01/26/2021 19:08	4820250
CCL	11CI-PF3OUdS/F-53B Minor	537.1	2.0	--		2.1413	2.0	ng/L	107	50 - 150	---	---	1.0	01/26/2021 08:41	01/26/2021 19:08	4820250
CCL	Perfluorotetradecanoic acid (PFTeDA)	537.1	2.0	--		2.0107	2.0	ng/L	101	50 - 150	---	---	1.0	01/26/2021 08:41	01/26/2021 19:08	4820250
CCL	SS-HFO-DA-13C3	537.1	N/A	--		41.5678	40.0	ng/L	104	70 - 130	---	---	1.0	01/26/2021 08:41	01/26/2021 19:08	4820250
LRB	Perfluorooctanoic acid (PFOA)	537.1	2.0	--	<	2.0		ng/L	---	---	---	---	1.0	01/26/2021 06:20	01/26/2021 19:34	4820237
LRB	Perfluorooctanesulfonic acid (PFOS)	537.1	2.0	--	<	2.0		ng/L	---	---	---	---	1.0	01/26/2021 06:20	01/26/2021 19:34	4820237
LRB	IS-NMeFOSAA-d3	537.1	N/A	--		183061	174812	ng/L	105	50 - 150	---	---	1.0	01/26/2021 06:20	01/26/2021 19:34	4820237
LRB	IS-PFOA-13C2	537.1	N/A	--		512977	449813	ng/L	114	50 - 150	---	---	1.0	01/26/2021 06:20	01/26/2021 19:34	4820237
LRB	IS-PFOS-13C4	537.1	N/A	--		245086	227165	ng/L	108	50 - 150	---	---	1.0	01/26/2021 06:20	01/26/2021 19:34	4820237
LRB	SS-NEFOSAA-d5	537.1	N/A	--		137.0090	160	ng/L	86	70 - 130	---	---	1.0	01/26/2021 06:20	01/26/2021 19:34	4820237
LRB	SS-PFDA-13C2	537.1	N/A	--		34.1094	40.0	ng/L	85	70 - 130	---	---	1.0	01/26/2021 06:20	01/26/2021 19:34	4820237
LRB	SS-PFHXA-13C2	537.1	N/A	--		35.7593	40.0	ng/L	89	70 - 130	---	---	1.0	01/26/2021 06:20	01/26/2021 19:34	4820237
LRB	Perfluorobutanesulfonic acid (PFBS)	537.1	2.0	--	<	2.0		ng/L	---	---	---	---	1.0	01/26/2021 06:20	01/26/2021 19:34	4820237
LRB	Perfluorheptanoic acid (PFHpA)	537.1	2.0	--	<	2.0		ng/L	---	---	---	---	1.0	01/26/2021 06:20	01/26/2021 19:34	4820237
LRB	Perfluorohexanesulfonic acid (PFHxS)	537.1	2.0	--	<	2.0		ng/L	---	---	---	---	1.0	01/26/2021 06:20	01/26/2021 19:34	4820237
LRB	Perfluorononanoic acid (PFNA)	537.1	2.0	--	<	2.0		ng/L	---	---	---	---	1.0	01/26/2021 06:20	01/26/2021 19:34	4820237
LRB	Perfluorodecanoic acid (PFDA)	537.1	2.0	--	<	2.0		ng/L	---	---	---	---	1.0	01/26/2021 06:20	01/26/2021 19:34	4820237
LRB	Perfluorohexanoic acid (PFHxA)	537.1	2.0	--	<	2.0		ng/L	---	---	---	---	1.0	01/26/2021 06:20	01/26/2021 19:34	4820237
LRB	Perfluorododecanoic acid (PFDoA)	537.1	2.0	--	<	2.0		ng/L	---	---	---	---	1.0	01/26/2021 06:20	01/26/2021 19:34	4820237

QC Summary Report (cont.)

Sample Type	Analyte	Method	MRL	Client ID	Result Flag	Amount	Target	Units	% Recovery	Recovery Limits	RPD	RPD Limit	Dil Factor	Extracted	Analyzed	EEA ID #
LRB	Perfluorotridecanoic acid (PFTrDA)	537.1	2.0	---	<	2.0		ng/L	---	---	---	---	1.0	01/26/2021 06:20	01/26/2021 19:34	4820237
LRB	Perfluoroundecanoic acid (PFUnA)	537.1	2.0	---	<	2.0		ng/L	---	---	---	---	1.0	01/26/2021 06:20	01/26/2021 19:34	4820237
LRB	N-ethyl Perfluorooctanesulfonamidoacetic acid	537.1	2.0	---	<	2.0		ng/L	---	---	---	---	1.0	01/26/2021 06:20	01/26/2021 19:34	4820237
LRB	N-methyl Perfluorooctanesulfonamidoacetic acid	537.1	2.0	---	<	2.0		ng/L	---	---	---	---	1.0	01/26/2021 06:20	01/26/2021 19:34	4820237
LRB	HFPO-DA/GenX	537.1	2.0	---	<	2.0		ng/L	---	---	---	---	1.0	01/26/2021 06:20	01/26/2021 19:34	4820237
LRB	ADONA	537.1	2.0	---	<	2.0		ng/L	---	---	---	---	1.0	01/26/2021 06:20	01/26/2021 19:34	4820237
LRB	9CI-PF3ONS/F-53B Major	537.1	2.0	---	<	2.0		ng/L	---	---	---	---	1.0	01/26/2021 06:20	01/26/2021 19:34	4820237
LRB	11CI-PF3OUdS/F-53B Minor	537.1	2.0	---	<	2.0		ng/L	---	---	---	---	1.0	01/26/2021 06:20	01/26/2021 19:34	4820237
LRB	Perfluorotetradecanoic acid (PFTeDA)	537.1	2.0	---	<	2.0		ng/L	---	---	---	---	1.0	01/26/2021 06:20	01/26/2021 19:34	4820237
LRB	SS-HFPO-DA-13C3	537.1	N/A	---		34.1677	40.0	ng/L	85	70 - 130	---	---	1.0	01/26/2021 06:20	01/26/2021 19:34	4820237
FBL	Perfluorooctanoic acid (PFOA)	537.1	2.0	---		1.6715	2.0	ng/L	84	50 - 150	---	---	1.0	01/26/2021 06:20	01/26/2021 19:47	4820238
FBL	Perfluorooctanesulfonic acid (PFOS)	537.1	2.0	---		1.7070	2.0	ng/L	85	50 - 150	---	---	1.0	01/26/2021 06:20	01/26/2021 19:47	4820238
FBL	IS-NMeFOSAA-d3	537.1	N/A	---		181828	174812	ng/L	104	50 - 150	---	---	1.0	01/26/2021 06:20	01/26/2021 19:47	4820238
FBL	IS-PFOA-13C2	537.1	N/A	---		502628	449813	ng/L	112	50 - 150	---	---	1.0	01/26/2021 06:20	01/26/2021 19:47	4820238
FBL	IS-PFOS-13C4	537.1	N/A	---		239312	227165	ng/L	105	50 - 150	---	---	1.0	01/26/2021 06:20	01/26/2021 19:47	4820238
FBL	SS-NEIFOSAA-d5	537.1	N/A	---		144.8230	160	ng/L	91	70 - 130	---	---	1.0	01/26/2021 06:20	01/26/2021 19:47	4820238
FBL	SS-PFDA-13C2	537.1	N/A	---		35.7256	40.0	ng/L	89	70 - 130	---	---	1.0	01/26/2021 06:20	01/26/2021 19:47	4820238
FBL	SS-PFHXA-13C2	537.1	N/A	---		35.4249	40.0	ng/L	89	70 - 130	---	---	1.0	01/26/2021 06:20	01/26/2021 19:47	4820238
FBL	Perfluorobutanesulfonic acid (PFBS)	537.1	2.0	---		1.2217	2.0	ng/L	61	50 - 150	---	---	1.0	01/26/2021 06:20	01/26/2021 19:47	4820238
FBL	Perfluorohexanesulfonic acid (PFHxS)	537.1	2.0	---		1.4618	2.0	ng/L	73	50 - 150	---	---	1.0	01/26/2021 06:20	01/26/2021 19:47	4820238
FBL	Perfluorooctanesulfonic acid (PFHxS)	537.1	2.0	---		1.4808	2.0	ng/L	74	50 - 150	---	---	1.0	01/26/2021 06:20	01/26/2021 19:47	4820238
FBL	Perfluorononanoic acid (PFNA)	537.1	2.0	---		1.5405	2.0	ng/L	77	50 - 150	---	---	1.0	01/26/2021 06:20	01/26/2021 19:47	4820238
FBL	Perfluorodecanoic acid (PFDA)	537.1	2.0	---		1.4018	2.0	ng/L	70	50 - 150	---	---	1.0	01/26/2021 06:20	01/26/2021 19:47	4820238
FBL	Perfluorohexanoic acid (PFHxA)	537.1	2.0	---		1.4026	2.0	ng/L	70	50 - 150	---	---	1.0	01/26/2021 06:20	01/26/2021 19:47	4820238
FBL	Perfluorododecanoic acid (PFDoA)	537.1	2.0	---		1.4573	2.0	ng/L	73	50 - 150	---	---	1.0	01/26/2021 06:20	01/26/2021 19:47	4820238
FBL	Perfluorotridecanoic acid (PFTrDA)	537.1	2.0	---		1.3915	2.0	ng/L	70	50 - 150	---	---	1.0	01/26/2021 06:20	01/26/2021 19:47	4820238
FBL	Perfluoroundecanoic acid (PFUnA)	537.1	2.0	---		1.4344	2.0	ng/L	72	50 - 150	---	---	1.0	01/26/2021 06:20	01/26/2021 19:47	4820238
FBL	N-ethyl Perfluorooctanesulfonamidoacetic acid	537.1	2.0	---		1.3771	2.0	ng/L	69	50 - 150	---	---	1.0	01/26/2021 06:20	01/26/2021 19:47	4820238
FBL	N-methyl Perfluorooctanesulfonamidoacetic acid	537.1	2.0	---		1.4089	2.0	ng/L	70	50 - 150	---	---	1.0	01/26/2021 06:20	01/26/2021 19:47	4820238
FBL	HFPO-DA/GenX	537.1	2.0	---		1.2976	2.0	ng/L	65	50 - 150	---	---	1.0	01/26/2021 06:20	01/26/2021 19:47	4820238
FBL	ADONA	537.1	2.0	---		1.2745	2.0	ng/L	64	50 - 150	---	---	1.0	01/26/2021 06:20	01/26/2021 19:47	4820238
FBL	9CI-PF3ONS/F-53B Major	537.1	2.0	---		1.5996	2.0	ng/L	80	50 - 150	---	---	1.0	01/26/2021 06:20	01/26/2021 19:47	4820238
FBL	11CI-PF3OUdS/F-53B Minor	537.1	2.0	---		1.6108	2.0	ng/L	81	50 - 150	---	---	1.0	01/26/2021 06:20	01/26/2021 19:47	4820238
FBL	Perfluorotetradecanoic acid (PFTeDA)	537.1	2.0	---		1.5262	2.0	ng/L	76	50 - 150	---	---	1.0	01/26/2021 06:20	01/26/2021 19:47	4820238
FBL	SS-HFPO-DA-13C3	537.1	N/A	---		35.5323	40.0	ng/L	89	70 - 130	---	---	1.0	01/26/2021 06:20	01/26/2021 19:47	4820238
FBL	Perfluorooctanoic acid (PFOA)	537.1	2.0	---		195.3890	200	ng/L	98	70 - 130	---	---	1.0	01/26/2021 06:20	01/26/2021 20:00	4820241
FBL	Perfluorooctanesulfonic acid (PFOS)	537.1	2.0	---		199.3710	200	ng/L	100	70 - 130	---	---	1.0	01/26/2021 06:20	01/26/2021 20:00	4820241
FBL	IS-NMeFOSAA-d3	537.1	N/A	---		187781	174812	ng/L	107	50 - 150	---	---	1.0	01/26/2021 06:20	01/26/2021 20:00	4820241
FBL	IS-PFOA-13C2	537.1	N/A	---		498704	449813	ng/L	111	50 - 150	---	---	1.0	01/26/2021 06:20	01/26/2021 20:00	4820241
FBL	IS-PFOS-13C4	537.1	N/A	---		233652	227165	ng/L	103	50 - 150	---	---	1.0	01/26/2021 06:20	01/26/2021 20:00	4820241
FBL	SS-NEIFOSAA-d5	537.1	N/A	---		145.6280	160	ng/L	91	70 - 130	---	---	1.0	01/26/2021 06:20	01/26/2021 20:00	4820241

QC Summary Report (cont.)

Sample Type	Analyte	Method	MRL	Client ID	Result Flag	Amount	Target	Units	% Recovery	Recovery Limits	RPD	RPD Limit	Dil Factor	Extracted	Analyzed	EEA ID #
FBH	SS-PFDA-13C2	537.1	N/A	---		37.9793	40.0	ng/L	95	70 - 130	---	---	1.0	01/26/2021 06:20	01/26/2021 20:00	4820241
FBH	SS-PFHXA-13C2	537.1	N/A	---		38.8216	40.0	ng/L	97	70 - 130	---	---	1.0	01/26/2021 06:20	01/26/2021 20:00	4820241
FBH	Perfluorobutanesulfonic acid (PFBS)	537.1	2.0	---		192.5520	200	ng/L	96	70 - 130	---	---	1.0	01/26/2021 06:20	01/26/2021 20:00	4820241
FBH	Perfluoroheptanoic acid (PFHpA)	537.1	2.0	---		191.6770	200	ng/L	96	70 - 130	---	---	1.0	01/26/2021 06:20	01/26/2021 20:00	4820241
FBH	Perfluorohexanesulfonic acid (PFHxS)	537.1	2.0	---		204.2850	200	ng/L	102	70 - 130	---	---	1.0	01/26/2021 06:20	01/26/2021 20:00	4820241
FBH	Perfluorononanoic acid (PFNA)	537.1	2.0	---		186.1500	200	ng/L	93	70 - 130	---	---	1.0	01/26/2021 06:20	01/26/2021 20:00	4820241
FBH	Perfluorodecanoic acid (PFDA)	537.1	2.0	---		181.2680	200	ng/L	91	70 - 130	---	---	1.0	01/26/2021 06:20	01/26/2021 20:00	4820241
FBH	Perfluorohexanoic acid (PFHxA)	537.1	2.0	---		183.5470	200	ng/L	92	70 - 130	---	---	1.0	01/26/2021 06:20	01/26/2021 20:00	4820241
FBH	Perfluorododecanoic acid (PFDoA)	537.1	2.0	---		176.5270	200	ng/L	88	70 - 130	---	---	1.0	01/26/2021 06:20	01/26/2021 20:00	4820241
FBH	Perfluorotridecanoic acid (PFTTDA)	537.1	2.0	---		180.6530	200	ng/L	90	70 - 130	---	---	1.0	01/26/2021 06:20	01/26/2021 20:00	4820241
FBH	Perfluoroundecanoic acid (PFUnA)	537.1	2.0	---		179.3940	200	ng/L	90	70 - 130	---	---	1.0	01/26/2021 06:20	01/26/2021 20:00	4820241
FBH	N-ethyl Perfluorooctanesulfonamidoacetic acid	537.1	2.0	---		184.8390	200	ng/L	92	70 - 130	---	---	1.0	01/26/2021 06:20	01/26/2021 20:00	4820241
FBH	N-methyl Perfluorooctanesulfonamidoacetic acid	537.1	2.0	---		189.9190	200	ng/L	95	70 - 130	---	---	1.0	01/26/2021 06:20	01/26/2021 20:00	4820241
FBH	HFPO-DA/GenX	537.1	2.0	---		179.6050	200	ng/L	90	70 - 130	---	---	1.0	01/26/2021 06:20	01/26/2021 20:00	4820241
FBH	ADONA	537.1	2.0	---		191.0850	200	ng/L	96	70 - 130	---	---	1.0	01/26/2021 06:20	01/26/2021 20:00	4820241
FBH	9C+PF3ONS/F-53B Major	537.1	2.0	---		190.4950	200	ng/L	95	70 - 130	---	---	1.0	01/26/2021 06:20	01/26/2021 20:00	4820241
FBH	11C+PF3OUs/F-53B Minor	537.1	2.0	---		187.3320	200	ng/L	94	70 - 130	---	---	1.0	01/26/2021 06:20	01/26/2021 20:00	4820241
FBH	Perfluorotetradecanoic acid (PFTeDA)	537.1	2.0	---		196.9240	200	ng/L	98	70 - 130	---	---	1.0	01/26/2021 06:20	01/26/2021 20:00	4820241
FBH	SS-HFPO-DA-13C3	537.1	N/A	---		39.2514	40.0	ng/L	98	70 - 130	---	---	1.0	01/26/2021 06:20	01/26/2021 20:00	4820241
CCM	Perfluorooctanoic acid (PFOA)	537.1	2.0	---		100.3560	100	ng/L	100	70 - 130	---	---	1.0	01/26/2021 08:41	01/27/2021 00:21	4820251
CCM	Perfluorooctanesulfonic acid (PFOS)	537.1	2.0	---		100.0210	100	ng/L	100	70 - 130	---	---	1.0	01/26/2021 08:41	01/27/2021 00:21	4820251
CCM	IS-NMeFOSAA-d3	537.1	N/A	---		162884	162884	ng/L	100	50 - 150	---	---	1.0	01/26/2021 08:41	01/27/2021 00:21	4820251
CCM	IS-PFOA-13C2	537.1	N/A	---		414119	414119	ng/L	100	50 - 150	---	---	1.0	01/26/2021 08:41	01/27/2021 00:21	4820251
CCM	IS-PFOS-13C4	537.1	N/A	---		207547	207547	ng/L	100	50 - 150	---	---	1.0	01/26/2021 08:41	01/27/2021 00:21	4820251
CCM	SS-NEFOSAA-d5	537.1	N/A	---		156.8460	160	ng/L	98	70 - 130	---	---	1.0	01/26/2021 08:41	01/27/2021 00:21	4820251
CCM	SS-PFDA-13C2	537.1	N/A	---		39.6612	40.0	ng/L	99	70 - 130	---	---	1.0	01/26/2021 08:41	01/27/2021 00:21	4820251
CCM	SS-PFHXA-13C2	537.1	N/A	---		43.0735	40.0	ng/L	108	70 - 130	---	---	1.0	01/26/2021 08:41	01/27/2021 00:21	4820251
CCM	Perfluorobutanesulfonic acid (PFBS)	537.1	2.0	---		102.1670	100	ng/L	102	70 - 130	---	---	1.0	01/26/2021 08:41	01/27/2021 00:21	4820251
CCM	Perfluoroheptanoic acid (PFHpA)	537.1	2.0	---		102.6440	100	ng/L	103	70 - 130	---	---	1.0	01/26/2021 08:41	01/27/2021 00:21	4820251
CCM	Perfluorohexanesulfonic acid (PFHxS)	537.1	2.0	---		102.4900	100	ng/L	102	70 - 130	---	---	1.0	01/26/2021 08:41	01/27/2021 00:21	4820251
CCM	Perfluorononanoic acid (PFNA)	537.1	2.0	---		99.9964	100	ng/L	100	70 - 130	---	---	1.0	01/26/2021 08:41	01/27/2021 00:21	4820251
CCM	Perfluorodecanoic acid (PFDA)	537.1	2.0	---		97.4064	100	ng/L	97	70 - 130	---	---	1.0	01/26/2021 08:41	01/27/2021 00:21	4820251
CCM	Perfluorohexanoic acid (PFHxA)	537.1	2.0	---		110.2790	100	ng/L	110	70 - 130	---	---	1.0	01/26/2021 08:41	01/27/2021 00:21	4820251
CCM	Perfluorododecanoic acid (PFDoA)	537.1	2.0	---		103.6600	100	ng/L	104	70 - 130	---	---	1.0	01/26/2021 08:41	01/27/2021 00:21	4820251
CCM	Perfluorotridecanoic acid (PFTTDA)	537.1	2.0	---		104.1370	100	ng/L	104	70 - 130	---	---	1.0	01/26/2021 08:41	01/27/2021 00:21	4820251
CCM	Perfluoroundecanoic acid (PFUnA)	537.1	2.0	---		101.7940	100	ng/L	102	70 - 130	---	---	1.0	01/26/2021 08:41	01/27/2021 00:21	4820251
CCM	N-ethyl Perfluorooctanesulfonamidoacetic acid	537.1	2.0	---		101.5870	100	ng/L	102	70 - 130	---	---	1.0	01/26/2021 08:41	01/27/2021 00:21	4820251
CCM	N-methyl Perfluorooctanesulfonamidoacetic acid	537.1	2.0	---		101.5430	100	ng/L	102	70 - 130	---	---	1.0	01/26/2021 08:41	01/27/2021 00:21	4820251
CCM	ADONA	537.1	2.0	---		105.1610	100	ng/L	105	70 - 130	---	---	1.0	01/26/2021 08:41	01/27/2021 00:21	4820251
CCM	9C+PF3ONS/F-53B Major	537.1	2.0	---		100.0920	100	ng/L	100	70 - 130	---	---	1.0	01/26/2021 08:41	01/27/2021 00:21	4820251
CCM	9C+PF3ONS/F-53B Major	537.1	2.0	---		101.1940	100	ng/L	101	70 - 130	---	---	1.0	01/26/2021 08:41	01/27/2021 00:21	4820251

QC Summary Report (cont.)

Sample Type	Analyte	Method	MRL	Client ID	Result Flag	Amount	Target	Units	% Recovery	Recovery Limits	RPD	RPD Limit	Dil Factor	Extracted	Analyzed	EEA ID #
CCM	11Cl-PF300dSF-53B Minor	537.1	2.0	---		102.0330	100	ng/L	102	70 - 130	---	---	1.0	01/26/2021 08:41	01/27/2021 00:21	4820251
CCM	Perfluorotetradecanoic acid (PFTeDA)	537.1	2.0	---		136.9440	100	ng/L	137	70 - 130	---	---	1.0	01/26/2021 08:41	01/27/2021 00:21	4820251
CCM	SS-HFPO-DA-13C3	537.1	N/A	---		46.2857	40.0	ng/L	116	70 - 130	---	---	1.0	01/26/2021 08:41	01/27/2021 00:21	4820251
CCH	Perfluorooctanoic acid (PFOA)	537.1	2.0	---		199.7930	200	ng/L	100	70 - 130	---	---	1.0	01/26/2021 08:41	01/27/2021 02:44	4820252
CCH	Perfluorooctanesulfonic acid (PFOS)	537.1	2.0	---		202.7870	200	ng/L	101	70 - 130	---	---	1.0	01/26/2021 08:41	01/27/2021 02:44	4820252
CCH	IS-NMeFOSAA-d3	537.1	N/A	---		168717	168717	ng/L	100	50 - 150	---	---	1.0	01/26/2021 08:41	01/27/2021 02:44	4820252
CCH	IS-PFOA-13C2	537.1	N/A	---		414580	414580	ng/L	100	50 - 150	---	---	1.0	01/26/2021 08:41	01/27/2021 02:44	4820252
CCH	IS-PFOS-13C4	537.1	N/A	---		209569	209569	ng/L	100	50 - 150	---	---	1.0	01/26/2021 08:41	01/27/2021 02:44	4820252
CCH	SS-NEFOSAA-d5	537.1	N/A	---		153.8470	160	ng/L	96	70 - 130	---	---	1.0	01/26/2021 08:41	01/27/2021 02:44	4820252
CCH	SS-PFDA-13C2	537.1	N/A	---		42.2673	40.0	ng/L	106	70 - 130	---	---	1.0	01/26/2021 08:41	01/27/2021 02:44	4820252
CCH	SS-PFHXA-13C2	537.1	N/A	---		43.9882	40.0	ng/L	110	70 - 130	---	---	1.0	01/26/2021 08:41	01/27/2021 02:44	4820252
CCH	Perfluorobutanesulfonic acid (PFBS)	537.1	2.0	---		203.2120	200	ng/L	102	70 - 130	---	---	1.0	01/26/2021 08:41	01/27/2021 02:44	4820252
CCH	Perfluorohexanoic acid (PFHxA)	537.1	2.0	---		206.4270	200	ng/L	103	70 - 130	---	---	1.0	01/26/2021 08:41	01/27/2021 02:44	4820252
CCH	Perfluorohexanesulfonic acid (PFHxS)	537.1	2.0	---		205.9270	200	ng/L	103	70 - 130	---	---	1.0	01/26/2021 08:41	01/27/2021 02:44	4820252
CCH	Perfluorononanoic acid (PFNA)	537.1	2.0	---		202.0290	200	ng/L	101	70 - 130	---	---	1.0	01/26/2021 08:41	01/27/2021 02:44	4820252
CCH	Perfluorodecanoic acid (PFDA)	537.1	2.0	---		199.1490	200	ng/L	100	70 - 130	---	---	1.0	01/26/2021 08:41	01/27/2021 02:44	4820252
CCH	Perfluorohexanoic acid (PFHxA)	537.1	2.0	---		217.4950	200	ng/L	109	70 - 130	---	---	1.0	01/26/2021 08:41	01/27/2021 02:44	4820252
CCH	Perfluorododecanoic acid (PFDoA)	537.1	2.0	---		204.9790	200	ng/L	102	70 - 130	---	---	1.0	01/26/2021 08:41	01/27/2021 02:44	4820252
CCH	Perfluorotridecanoic acid (PFTriDA)	537.1	2.0	---		211.5870	200	ng/L	106	70 - 130	---	---	1.0	01/26/2021 08:41	01/27/2021 02:44	4820252
CCH	Perfluoroundecanoic acid (PFUnA)	537.1	2.0	---		197.5930	200	ng/L	99	70 - 130	---	---	1.0	01/26/2021 08:41	01/27/2021 02:44	4820252
CCH	N-methyl Perfluorooctanesulfonamidoacetic acid	537.1	2.0	---		197.1310	200	ng/L	99	70 - 130	---	---	1.0	01/26/2021 08:41	01/27/2021 02:44	4820252
CCH	N-methyl Perfluorooctanesulfonamidoacetic acid	537.1	2.0	---		203.5030	200	ng/L	102	70 - 130	---	---	1.0	01/26/2021 08:41	01/27/2021 02:44	4820252
CCH	HFPO-DA/GenX	537.1	2.0	---		216.0690	200	ng/L	108	70 - 130	---	---	1.0	01/26/2021 08:41	01/27/2021 02:44	4820252
CCH	ADONA	537.1	2.0	---		214.2180	200	ng/L	107	70 - 130	---	---	1.0	01/26/2021 08:41	01/27/2021 02:44	4820252
CCH	9Cl-PF3ONSIF-53B Major	537.1	2.0	---		193.4990	200	ng/L	97	70 - 130	---	---	1.0	01/26/2021 08:41	01/27/2021 02:44	4820252
CCH	11Cl-PF300dSF-53B Minor	537.1	2.0	---		196.8880	200	ng/L	98	70 - 130	---	---	1.0	01/26/2021 08:41	01/27/2021 02:44	4820252
CCH	Perfluorotetradecanoic acid (PFTeDA)	537.1	2.0	---		253.3250	200	ng/L	127	70 - 130	---	---	1.0	01/26/2021 08:41	01/27/2021 02:44	4820252
CCH	SS-HFPO-DA-13C3	537.1	N/A	---		45.5317	40.0	ng/L	114	70 - 130	---	---	1.0	01/26/2021 08:41	01/27/2021 02:44	4820252
CCM	Perfluorooctanoic acid (PFOA)	537.1	2.0	---		101.7880	100	ng/L	102	70 - 130	---	---	1.0	01/26/2021 08:41	01/27/2021 10:08	4820921
CCM	Perfluorooctanesulfonic acid (PFOS)	537.1	2.0	---		99.3150	100	ng/L	99	70 - 130	---	---	1.0	01/26/2021 08:41	01/27/2021 10:08	4820921
CCM	IS-NMeFOSAA-d3	537.1	N/A	---		164962	164962	ng/L	100	50 - 150	---	---	1.0	01/26/2021 08:41	01/27/2021 10:08	4820921
CCM	IS-PFOA-13C2	537.1	N/A	---		419557	419557	ng/L	100	50 - 150	---	---	1.0	01/26/2021 08:41	01/27/2021 10:08	4820921
CCM	IS-PFOS-13C4	537.1	N/A	---		218824	218824	ng/L	100	50 - 150	---	---	1.0	01/26/2021 08:41	01/27/2021 10:08	4820921
CCM	SS-NEFOSAA-d5	537.1	N/A	---		163.8630	160	ng/L	102	70 - 130	---	---	1.0	01/26/2021 08:41	01/27/2021 10:08	4820921
CCM	SS-PFDA-13C2	537.1	N/A	---		41.3071	40.0	ng/L	103	70 - 130	---	---	1.0	01/26/2021 08:41	01/27/2021 10:08	4820921
CCM	SS-PFHXA-13C2	537.1	N/A	---		43.9725	40.0	ng/L	110	70 - 130	---	---	1.0	01/26/2021 08:41	01/27/2021 10:08	4820921
CCM	Perfluorobutanesulfonic acid (PFBS)	537.1	2.0	---		103.7870	100	ng/L	104	70 - 130	---	---	1.0	01/26/2021 08:41	01/27/2021 10:08	4820921
CCM	Perfluorohexanoic acid (PFHxA)	537.1	2.0	---		101.8460	100	ng/L	102	70 - 130	---	---	1.0	01/26/2021 08:41	01/27/2021 10:08	4820921
CCM	Perfluorohexanesulfonic acid (PFHxS)	537.1	2.0	---		97.9714	100	ng/L	98	70 - 130	---	---	1.0	01/26/2021 08:41	01/27/2021 10:08	4820921
CCM	Perfluorononanoic acid (PFNA)	537.1	2.0	---		102.1940	100	ng/L	102	70 - 130	---	---	1.0	01/26/2021 08:41	01/27/2021 10:08	4820921
CCM	Perfluorodecanoic acid (PFDA)	537.1	2.0	---		102.4500	100	ng/L	102	70 - 130	---	---	1.0	01/26/2021 08:41	01/27/2021 10:08	4820921

QC Summary Report (cont.)

Sample Type	Analyte	Method	MRL	Client ID	Result Flag	Amount	Target	Units	% Recovery	Recovery Limits	RPD	RPD Limit	Dil Factor	Extracted	Analyzed	EEA ID #
CCM	Perfluorohexanoic acid (PFHxA)	537.1	2.0	---		110.0410	100	ng/L	110	70 - 130	---	---	1.0	01/26/2021 08:41	01/27/2021 10:08	4820921
CCM	Perfluorododecanoic acid (PFDoA)	537.1	2.0	---		103.5940	100	ng/L	104	70 - 130	---	---	1.0	01/26/2021 08:41	01/27/2021 10:08	4820921
CCM	Perfluorotridecanoic acid (PFTriDA)	537.1	2.0	---		103.6050	100	ng/L	104	70 - 130	---	---	1.0	01/26/2021 08:41	01/27/2021 10:08	4820921
CCM	Perfluoroundecanoic acid (PFUnA)	537.1	2.0	---		103.9880	100	ng/L	104	70 - 130	---	---	1.0	01/26/2021 08:41	01/27/2021 10:08	4820921
CCM	N-ethyl Perfluorooctanesulfonamidoacetic acid	537.1	2.0	---		106.1150	100	ng/L	106	70 - 130	---	---	1.0	01/26/2021 08:41	01/27/2021 10:08	4820921
CCM	N-methyl Perfluorooctanesulfonamidoacetic acid	537.1	2.0	---		101.7910	100	ng/L	102	70 - 130	---	---	1.0	01/26/2021 08:41	01/27/2021 10:08	4820921
CCM	HFPO-DA/GenX	537.1	2.0	---		106.2900	100	ng/L	106	70 - 130	---	---	1.0	01/26/2021 08:41	01/27/2021 10:08	4820921
CCM	ADONA	537.1	2.0	---		98.4254	100	ng/L	98	70 - 130	---	---	1.0	01/26/2021 08:41	01/27/2021 10:08	4820921
CCM	9CI-PF3ONS/F-53B Major	537.1	2.0	---		102.2820	100	ng/L	102	70 - 130	---	---	1.0	01/26/2021 08:41	01/27/2021 10:08	4820921
CCM	11CI-PF3OUdS/F-53B Minor	537.1	2.0	---		100.8660	100	ng/L	101	70 - 130	---	---	1.0	01/26/2021 08:41	01/27/2021 10:08	4820921
CCM	Perfluorotetradecanoic acid (PFTeDA)	537.1	2.0	---		110.9090	100	ng/L	111	70 - 130	---	---	1.0	01/26/2021 08:41	01/27/2021 10:08	4820921
CCM	SS-HFPO-DA-13C3	537.1	N/A	---		43.7745	40.0	ng/L	109	70 - 130	---	---	1.0	01/26/2021 08:41	01/27/2021 10:08	4820921
CCH	Perfluorooctanoic acid (PFOA)	537.1	2.0	---		200.3460	200	ng/L	100	70 - 130	---	---	1.0	01/26/2021 08:41	01/27/2021 11:20	4820920
CCH	Perfluorooctanesulfonic acid (PFOS)	537.1	2.0	---		202.2700	200	ng/L	101	70 - 130	---	---	1.0	01/26/2021 08:41	01/27/2021 11:20	4820920
CCH	IS-NMeFOSAA-d3	537.1	N/A	---		180631	180631	ng/L	100	50 - 150	---	---	1.0	01/26/2021 08:41	01/27/2021 11:20	4820920
CCH	IS-PFOA-13C2	537.1	N/A	---		426683	426683	ng/L	100	50 - 150	---	---	1.0	01/26/2021 08:41	01/27/2021 11:20	4820920
CCH	IS-PFOS-13C4	537.1	N/A	---		223104	223104	ng/L	100	50 - 150	---	---	1.0	01/26/2021 08:41	01/27/2021 11:20	4820920
CCH	SS-NEFOSAA-d5	537.1	N/A	---		157.6820	160	ng/L	99	70 - 130	---	---	1.0	01/26/2021 08:41	01/27/2021 11:20	4820920
CCH	SS-PFDA-13C2	537.1	N/A	---		41.6765	40.0	ng/L	104	70 - 130	---	---	1.0	01/26/2021 08:41	01/27/2021 11:20	4820920
CCH	SS-PFHxA-13C2	537.1	N/A	---		44.6445	40.0	ng/L	112	70 - 130	---	---	1.0	01/26/2021 08:41	01/27/2021 11:20	4820920
CCH	Perfluorobutanesulfonic acid (PFBS)	537.1	2.0	---		206.4410	200	ng/L	103	70 - 130	---	---	1.0	01/26/2021 08:41	01/27/2021 11:20	4820920
CCH	Perfluorohexanoic acid (PFHxA)	537.1	2.0	---		205.3630	200	ng/L	103	70 - 130	---	---	1.0	01/26/2021 08:41	01/27/2021 11:20	4820920
CCH	Perfluorohexanesulfonic acid (PFHxS)	537.1	2.0	---		201.5090	200	ng/L	101	70 - 130	---	---	1.0	01/26/2021 08:41	01/27/2021 11:20	4820920
CCH	Perfluorononanoic acid (PFNA)	537.1	2.0	---		200.7470	200	ng/L	100	70 - 130	---	---	1.0	01/26/2021 08:41	01/27/2021 11:20	4820920
CCH	Perfluorodecanoic acid (PFDA)	537.1	2.0	---		205.5490	200	ng/L	103	70 - 130	---	---	1.0	01/26/2021 08:41	01/27/2021 11:20	4820920
CCH	Perfluorohexanoic acid (PFHxA)	537.1	2.0	---		218.2170	200	ng/L	109	70 - 130	---	---	1.0	01/26/2021 08:41	01/27/2021 11:20	4820920
CCH	Perfluorododecanoic acid (PFDoA)	537.1	2.0	---		215.7340	200	ng/L	108	70 - 130	---	---	1.0	01/26/2021 08:41	01/27/2021 11:20	4820920
CCH	Perfluorotridecanoic acid (PFTriDA)	537.1	2.0	---		213.2740	200	ng/L	107	70 - 130	---	---	1.0	01/26/2021 08:41	01/27/2021 11:20	4820920
CCH	Perfluoroundecanoic acid (PFUnA)	537.1	2.0	---		212.5080	200	ng/L	106	70 - 130	---	---	1.0	01/26/2021 08:41	01/27/2021 11:20	4820920
CCH	N-ethyl Perfluorooctanesulfonamidoacetic acid	537.1	2.0	---		202.9060	200	ng/L	101	70 - 130	---	---	1.0	01/26/2021 08:41	01/27/2021 11:20	4820920
CCH	N-methyl Perfluorooctanesulfonamidoacetic acid	537.1	2.0	---		204.0170	200	ng/L	102	70 - 130	---	---	1.0	01/26/2021 08:41	01/27/2021 11:20	4820920
CCH	HFPO-DA/GenX	537.1	2.0	---		218.2350	200	ng/L	109	70 - 130	---	---	1.0	01/26/2021 08:41	01/27/2021 11:20	4820920
CCH	ADONA	537.1	2.0	---		213.6690	200	ng/L	107	70 - 130	---	---	1.0	01/26/2021 08:41	01/27/2021 11:20	4820920
CCH	9CI-PF3ONS/F-53B Major	537.1	2.0	---		195.8950	200	ng/L	98	70 - 130	---	---	1.0	01/26/2021 08:41	01/27/2021 11:20	4820920
CCH	11CI-PF3OUdS/F-53B Minor	537.1	2.0	---		197.8650	200	ng/L	99	70 - 130	---	---	1.0	01/26/2021 08:41	01/27/2021 11:20	4820920
CCH	Perfluorotetradecanoic acid (PFTeDA)	537.1	2.0	---		277.8130	200	ng/L	139	70 - 130	---	---	1.0	01/26/2021 08:41	01/27/2021 11:20	4820920
CCH	SS-HFPO-DA-13C3	537.1	N/A	---		45.1203	40.0	ng/L	113	70 - 130	---	---	1.0	01/26/2021 08:41	01/27/2021 11:20	4820920

Eurofins Eaton Analytical

Run Log

Run ID: 284867 Method: 537.1

<u>Type</u>	<u>Sample Id</u>	<u>Sample Site</u>	<u>Matrix</u>	<u>Instrument ID</u>	<u>Analysis Date</u>	<u>Calibration File</u>
CCL	4822370		OS	GA	01/28/2021 22:14	012821M537.1a.wiff
LRB	4822023		RW	GA	01/28/2021 22:35	012821M537.1a.wiff
FBL	4822024		RW	GA	01/28/2021 22:45	012821M537.1a.wiff
FBM	4822025		RW	GA	01/28/2021 23:07	012821M537.1a.wiff
FS	4820255	Plant 1 Tap-LV Waterworks	DW	GA	01/29/2021 02:28	012821M537.1a.wiff
FS	4820256	Plant 2 Tap-LV Waterworks	DW	GA	01/29/2021 02:39	012821M537.1a.wiff
CCM	4822377		OS	GA	01/29/2021 03:43	012821M537.1a.wiff

QC Summary Report

Sample Type	Analyte	Method	MDA95	Client ID	Result Flag	Amount	Target	Units	% Recovery	Recovery Limits	RPD	RPD Limit	Dil Factor	Extracted	Analyzed	EEA ID #
CCL	Perfluorooctanoic acid (PFOA)	537.1	2.0	--		1.9915	2.0	ng/L	100	50 - 150	---	---	1.0	01/27/2021 13:34	01/28/2021 22:14	4822370
CCL	Perfluorooctanesulfonic acid (PFOS)	537.1	2.0	--		1.9287	2.0	ng/L	96	50 - 150	---	---	1.0	01/27/2021 13:34	01/28/2021 22:14	4822370
CCL	IS-NMeFOSAA-d3	537.1	N/A	--		1185846	1185846.2E	ng/L	100	50 - 150	---	---	1.0	01/27/2021 13:34	01/28/2021 22:14	4822370
CCL	IS-PFOA-13C2	537.1	N/A	--		1522481	1522480.84	ng/L	100	50 - 150	---	---	1.0	01/27/2021 13:34	01/28/2021 22:14	4822370
CCL	IS-PFOS-13C4	537.1	N/A	--		6759097	3759097.0E	ng/L	100	50 - 150	---	---	1.0	01/27/2021 13:34	01/28/2021 22:14	4822370
CCL	SS-NEFOSAA-d5	537.1	N/A	--		160.2778	160	ng/L	100	70 - 130	---	---	1.0	01/27/2021 13:34	01/28/2021 22:14	4822370
CCL	SS-PFDA-13C2	537.1	N/A	--		38.9612	40.0	ng/L	97	70 - 130	---	---	1.0	01/27/2021 13:34	01/28/2021 22:14	4822370
CCL	SS-PFHXA-13C2	537.1	N/A	--		39.2281	40.0	ng/L	98	70 - 130	---	---	1.0	01/27/2021 13:34	01/28/2021 22:14	4822370
CCL	Perfluorobutanesulfonic acid (PFBS)	537.1	2.0	--		1.9238	2.0	ng/L	96	50 - 150	---	---	1.0	01/27/2021 13:34	01/28/2021 22:14	4822370
CCL	Perfluorheptanoic acid (PFHpA)	537.1	2.0	--		1.9541	2.0	ng/L	98	50 - 150	---	---	1.0	01/27/2021 13:34	01/28/2021 22:14	4822370
CCL	Perfluorohexanesulfonic acid (PFHxS)	537.1	2.0	--		1.7436	2.0	ng/L	87	50 - 150	---	---	1.0	01/27/2021 13:34	01/28/2021 22:14	4822370
CCL	Perfluorononanoic acid (PFNA)	537.1	2.0	--		1.9650	2.0	ng/L	98	50 - 150	---	---	1.0	01/27/2021 13:34	01/28/2021 22:14	4822370
CCL	Perfluorodecanoic acid (PFDA)	537.1	2.0	--		1.9258	2.0	ng/L	96	50 - 150	---	---	1.0	01/27/2021 13:34	01/28/2021 22:14	4822370
CCL	Perfluorohexanoic acid (PFHxA)	537.1	2.0	--		1.9632	2.0	ng/L	98	50 - 150	---	---	1.0	01/27/2021 13:34	01/28/2021 22:14	4822370
CCL	Perfluorododecanoic acid (PFDoA)	537.1	2.0	--		1.9192	2.0	ng/L	96	50 - 150	---	---	1.0	01/27/2021 13:34	01/28/2021 22:14	4822370
CCL	Perfluorotridecanoic acid (PFTtDA)	537.1	2.0	--		1.8389	2.0	ng/L	92	50 - 150	---	---	1.0	01/27/2021 13:34	01/28/2021 22:14	4822370
CCL	Perfluoroundecanoic acid (PFUnA)	537.1	2.0	--		1.9126	2.0	ng/L	96	50 - 150	---	---	1.0	01/27/2021 13:34	01/28/2021 22:14	4822370
CCL	N-ethyl Perfluorooctanesulfonamidoacetic acid	537.1	2.0	--		1.9648	2.0	ng/L	98	50 - 150	---	---	1.0	01/27/2021 13:34	01/28/2021 22:14	4822370
CCL	N-methyl Perfluorooctanesulfonamidoacetic acid	537.1	2.0	--		2.0642	2.0	ng/L	103	50 - 150	---	---	1.0	01/27/2021 13:34	01/28/2021 22:14	4822370
CCL	HFPO-DA/GenX	537.1	2.0	--		1.9182	2.0	ng/L	96	50 - 150	---	---	1.0	01/27/2021 13:34	01/28/2021 22:14	4822370
CCL	ADONA	537.1	2.0	--		1.8770	2.0	ng/L	94	50 - 150	---	---	1.0	01/27/2021 13:34	01/28/2021 22:14	4822370
CCL	9CI-PF3ONS/F-53B Major	537.1	2.0	--		1.8583	2.0	ng/L	93	50 - 150	---	---	1.0	01/27/2021 13:34	01/28/2021 22:14	4822370
CCL	11CI-PF3OUdS/F-53B Minor	537.1	2.0	--		1.8442	2.0	ng/L	92	50 - 150	---	---	1.0	01/27/2021 13:34	01/28/2021 22:14	4822370
CCL	Perfluorotetradecanoic acid (PFTtDA)	537.1	2.0	--		1.8191	2.0	ng/L	91	50 - 150	---	---	1.0	01/27/2021 13:34	01/28/2021 22:14	4822370
CCL	SS-HFO-DA-13C3	537.1	N/A	--		38.9927	40.0	ng/L	97	70 - 130	---	---	1.0	01/27/2021 13:34	01/28/2021 22:14	4822370
LRB	Perfluorooctanoic acid (PFOA)	537.1	2.0	--	<	2.0		ng/L	---	---	---	---	1.0	01/28/2021 07:57	01/28/2021 22:35	4822023
LRB	Perfluorooctanesulfonic acid (PFOS)	537.1	2.0	--	<	2.0		ng/L	---	---	---	---	1.0	01/28/2021 07:57	01/28/2021 22:35	4822023
LRB	IS-NMeFOSAA-d3	537.1	N/A	--		1202111	1185846.2E	ng/L	101	50 - 150	---	---	1.0	01/28/2021 07:57	01/28/2021 22:35	4822023
LRB	IS-PFOA-13C2	537.1	N/A	--		1584452	1522480.84	ng/L	104	50 - 150	---	---	1.0	01/28/2021 07:57	01/28/2021 22:35	4822023
LRB	IS-PFOS-13C4	537.1	N/A	--		6835834	3759097.0E	ng/L	101	50 - 150	---	---	1.0	01/28/2021 07:57	01/28/2021 22:35	4822023
LRB	SS-NEFOSAA-d5	537.1	N/A	--		149.9330	160	ng/L	94	70 - 130	---	---	1.0	01/28/2021 07:57	01/28/2021 22:35	4822023
LRB	SS-PFDA-13C2	537.1	N/A	--		37.2960	40.0	ng/L	93	70 - 130	---	---	1.0	01/28/2021 07:57	01/28/2021 22:35	4822023
LRB	SS-PFHXA-13C2	537.1	N/A	--		38.9617	40.0	ng/L	97	70 - 130	---	---	1.0	01/28/2021 07:57	01/28/2021 22:35	4822023
LRB	Perfluorobutanesulfonic acid (PFBS)	537.1	2.0	--	<	2.0		ng/L	---	---	---	---	1.0	01/28/2021 07:57	01/28/2021 22:35	4822023
LRB	Perfluorheptanoic acid (PFHpA)	537.1	2.0	--	<	2.0		ng/L	---	---	---	---	1.0	01/28/2021 07:57	01/28/2021 22:35	4822023
LRB	Perfluorohexanesulfonic acid (PFHxS)	537.1	2.0	--	<	2.0		ng/L	---	---	---	---	1.0	01/28/2021 07:57	01/28/2021 22:35	4822023
LRB	Perfluorononanoic acid (PFNA)	537.1	2.0	--	<	2.0		ng/L	---	---	---	---	1.0	01/28/2021 07:57	01/28/2021 22:35	4822023
LRB	Perfluorodecanoic acid (PFDA)	537.1	2.0	--	<	2.0		ng/L	---	---	---	---	1.0	01/28/2021 07:57	01/28/2021 22:35	4822023
LRB	Perfluorohexanoic acid (PFHxA)	537.1	2.0	--	<	2.0		ng/L	---	---	---	---	1.0	01/28/2021 07:57	01/28/2021 22:35	4822023
LRB	Perfluorododecanoic acid (PFDoA)	537.1	2.0	--	<	2.0		ng/L	---	---	---	---	1.0	01/28/2021 07:57	01/28/2021 22:35	4822023

QC Summary Report (cont.)

Sample Type	Analyte	Method	MRL	Client ID	Result Flag	Amount	Target	Units	% Recovery	Recovery Limits	RPD	RPD Limit	Dil Factor	Extracted	Analyzed	EEA ID #
LRB	Perfluorotridecanoic acid (PFTrDA)	537.1	2.0	---	<	2.0		ng/L	---	---	---	---	1.0	01/28/2021 07:57	01/28/2021 22:35	4822023
LRB	Perfluoroundecanoic acid (PFUnA)	537.1	2.0	---	<	2.0		ng/L	---	---	---	---	1.0	01/28/2021 07:57	01/28/2021 22:35	4822023
LRB	N-ethyl Perfluorooctanesulfonamidoacetic acid	537.1	2.0	---	<	2.0		ng/L	---	---	---	---	1.0	01/28/2021 07:57	01/28/2021 22:35	4822023
LRB	N-methyl Perfluorooctanesulfonamidoacetic acid	537.1	2.0	---	<	2.0		ng/L	---	---	---	---	1.0	01/28/2021 07:57	01/28/2021 22:35	4822023
LRB	HFPO-DA/GenX	537.1	2.0	---	<	2.0		ng/L	---	---	---	---	1.0	01/28/2021 07:57	01/28/2021 22:35	4822023
LRB	ADONA	537.1	2.0	---	<	2.0		ng/L	---	---	---	---	1.0	01/28/2021 07:57	01/28/2021 22:35	4822023
LRB	9CI-PF3ONS/F-53B Major	537.1	2.0	---	<	2.0		ng/L	---	---	---	---	1.0	01/28/2021 07:57	01/28/2021 22:35	4822023
LRB	11CI-PF3OUdS/F-53B Minor	537.1	2.0	---	<	2.0		ng/L	---	---	---	---	1.0	01/28/2021 07:57	01/28/2021 22:35	4822023
LRB	Perfluorotetradecanoic acid (PFTeDA)	537.1	2.0	---	<	2.0		ng/L	---	---	---	---	1.0	01/28/2021 07:57	01/28/2021 22:35	4822023
LRB	SS-HFPO-DA-13C3	537.1	N/A	---		37.6702	40.0	ng/L	94	70 - 130	---	---	1.0	01/28/2021 07:57	01/28/2021 22:35	4822023
FBL	Perfluorooctanoic acid (PFOA)	537.1	2.0	---		2.0061	2.0	ng/L	100	50 - 150	---	---	1.0	01/28/2021 07:57	01/28/2021 22:45	4822024
FBL	Perfluorooctanesulfonic acid (PFOS)	537.1	2.0	---		1.8582	2.0	ng/L	93	50 - 150	---	---	1.0	01/28/2021 07:57	01/28/2021 22:45	4822024
FBL	IS-NMeFOSAA-d3	537.1	N/A	---		1217660	1185846.2E	ng/L	103	50 - 150	---	---	1.0	01/28/2021 07:57	01/28/2021 22:45	4822024
FBL	IS-PFOA-13C2	537.1	N/A	---		1545807	1522480.84	ng/L	102	50 - 150	---	---	1.0	01/28/2021 07:57	01/28/2021 22:45	4822024
FBL	IS-PFOS-13C4	537.1	N/A	---		6853520	3759097.03	ng/L	101	50 - 150	---	---	1.0	01/28/2021 07:57	01/28/2021 22:45	4822024
FBL	SS-NEFOSAA-d5	537.1	N/A	---		129.9957	160	ng/L	81	70 - 130	---	---	1.0	01/28/2021 07:57	01/28/2021 22:45	4822024
FBL	SS-PFDA-13C2	537.1	N/A	---		36.8372	40.0	ng/L	92	70 - 130	---	---	1.0	01/28/2021 07:57	01/28/2021 22:45	4822024
FBL	SS-PFHXA-13C2	537.1	N/A	---		36.6835	40.0	ng/L	92	70 - 130	---	---	1.0	01/28/2021 07:57	01/28/2021 22:45	4822024
FBL	Perfluorobutanesulfonic acid (PFBS)	537.1	2.0	---		1.7024	2.0	ng/L	85	50 - 150	---	---	1.0	01/28/2021 07:57	01/28/2021 22:45	4822024
FBL	Perfluorohexanoic acid (PFHxA)	537.1	2.0	---		1.9340	2.0	ng/L	97	50 - 150	---	---	1.0	01/28/2021 07:57	01/28/2021 22:45	4822024
FBL	Perfluorohexanesulfonic acid (PFHxS)	537.1	2.0	---		1.8149	2.0	ng/L	91	50 - 150	---	---	1.0	01/28/2021 07:57	01/28/2021 22:45	4822024
FBL	Perfluorononanoic acid (PFNA)	537.1	2.0	---		1.8125	2.0	ng/L	91	50 - 150	---	---	1.0	01/28/2021 07:57	01/28/2021 22:45	4822024
FBL	Perfluorodecanoic acid (PFDA)	537.1	2.0	---		1.8496	2.0	ng/L	92	50 - 150	---	---	1.0	01/28/2021 07:57	01/28/2021 22:45	4822024
FBL	Perfluorohexanoic acid (PFHxA)	537.1	2.0	---		1.9140	2.0	ng/L	96	50 - 150	---	---	1.0	01/28/2021 07:57	01/28/2021 22:45	4822024
FBL	Perfluorododecanoic acid (PFDoA)	537.1	2.0	---		1.5422	2.0	ng/L	77	50 - 150	---	---	1.0	01/28/2021 07:57	01/28/2021 22:45	4822024
FBL	Perfluorotridecanoic acid (PFTrDA)	537.1	2.0	---		1.4779	2.0	ng/L	74	50 - 150	---	---	1.0	01/28/2021 07:57	01/28/2021 22:45	4822024
FBL	Perfluoroundecanoic acid (PFUnA)	537.1	2.0	---		1.7109	2.0	ng/L	86	50 - 150	---	---	1.0	01/28/2021 07:57	01/28/2021 22:45	4822024
FBL	N-ethyl Perfluorooctanesulfonamidoacetic acid	537.1	2.0	---		1.8007	2.0	ng/L	90	50 - 150	---	---	1.0	01/28/2021 07:57	01/28/2021 22:45	4822024
FBL	N-methyl Perfluorooctanesulfonamidoacetic acid	537.1	2.0	---		1.8050	2.0	ng/L	90	50 - 150	---	---	1.0	01/28/2021 07:57	01/28/2021 22:45	4822024
FBL	HFPO-DA/GenX	537.1	2.0	---		1.8401	2.0	ng/L	92	50 - 150	---	---	1.0	01/28/2021 07:57	01/28/2021 22:45	4822024
FBL	ADONA	537.1	2.0	---		1.7697	2.0	ng/L	88	50 - 150	---	---	1.0	01/28/2021 07:57	01/28/2021 22:45	4822024
FBL	9CI-PF3ONS/F-53B Major	537.1	2.0	---		1.7581	2.0	ng/L	88	50 - 150	---	---	1.0	01/28/2021 07:57	01/28/2021 22:45	4822024
FBL	11CI-PF3OUdS/F-53B Minor	537.1	2.0	---		1.5623	2.0	ng/L	78	50 - 150	---	---	1.0	01/28/2021 07:57	01/28/2021 22:45	4822024
FBL	Perfluorotetradecanoic acid (PFTeDA)	537.1	2.0	---		1.5240	2.0	ng/L	76	50 - 150	---	---	1.0	01/28/2021 07:57	01/28/2021 22:45	4822024
FBL	SS-HFPO-DA-13C3	537.1	N/A	---		35.9097	40.0	ng/L	90	70 - 130	---	---	1.0	01/28/2021 07:57	01/28/2021 22:45	4822024
FBM	Perfluorooctanoic acid (PFOA)	537.1	2.0	---		99.1877	100	ng/L	99	70 - 130	---	---	1.0	01/28/2021 07:57	01/28/2021 23:07	4822025
FBM	Perfluorooctanesulfonic acid (PFOS)	537.1	2.0	---		98.9050	100	ng/L	99	70 - 130	---	---	1.0	01/28/2021 07:57	01/28/2021 23:07	4822025
FBM	IS-NMeFOSAA-d3	537.1	N/A	---		1234672	1185846.2E	ng/L	104	50 - 150	---	---	1.0	01/28/2021 07:57	01/28/2021 23:07	4822025
FBM	IS-PFOA-13C2	537.1	N/A	---		1614262	1522480.84	ng/L	106	50 - 150	---	---	1.0	01/28/2021 07:57	01/28/2021 23:07	4822025
FBM	IS-PFOS-13C4	537.1	N/A	---		6694636	3759097.03	ng/L	99	50 - 150	---	---	1.0	01/28/2021 07:57	01/28/2021 23:07	4822025
FBM	SS-NEFOSAA-d5	537.1	N/A	---		144.0596	160	ng/L	90	70 - 130	---	---	1.0	01/28/2021 07:57	01/28/2021 23:07	4822025

QC Summary Report (cont.)

Sample Type	Analyte	Method	MRL	Client ID	Result Flag	Amount	Target	Units	% Recovery	Recovery Limits	RPD	RPD Limit	Dil Factor	Extracted	Analyzed	EEA ID #
FBM	SS-PFDA-13C2	537.1	N/A	---		37.0939	40.0	ng/L	93	70 - 130	---	---	1.0	01/28/2021 07:57	01/28/2021 23:07	4822025
FBM	SS-PFHXA-13C2	537.1	N/A	---		38.2076	40.0	ng/L	96	70 - 130	---	---	1.0	01/28/2021 07:57	01/28/2021 23:07	4822025
FBM	Perfluorobutanesulfonic acid (PFBS)	537.1	2.0	---		100.3011	100	ng/L	100	70 - 130	---	---	1.0	01/28/2021 07:57	01/28/2021 23:07	4822025
FBM	Perfluoroheptanoic acid (PFHpA)	537.1	2.0	---		97.8860	100	ng/L	98	70 - 130	---	---	1.0	01/28/2021 07:57	01/28/2021 23:07	4822025
FBM	Perfluorohexanesulfonic acid (PFHxS)	537.1	2.0	---		99.4729	100	ng/L	99	70 - 130	---	---	1.0	01/28/2021 07:57	01/28/2021 23:07	4822025
FBM	Perfluorononanoic acid (PFNA)	537.1	2.0	---		98.8777	100	ng/L	99	70 - 130	---	---	1.0	01/28/2021 07:57	01/28/2021 23:07	4822025
FBM	Perfluorodecanoic acid (PFDA)	537.1	2.0	---		94.4348	100	ng/L	94	70 - 130	---	---	1.0	01/28/2021 07:57	01/28/2021 23:07	4822025
FBM	Perfluorohexanoic acid (PFHxA)	537.1	2.0	---		98.8384	100	ng/L	99	70 - 130	---	---	1.0	01/28/2021 07:57	01/28/2021 23:07	4822025
FBM	Perfluorododecanoic acid (PFDoA)	537.1	2.0	---		95.1279	100	ng/L	95	70 - 130	---	---	1.0	01/28/2021 07:57	01/28/2021 23:07	4822025
FBM	Perfluorotridecanoic acid (PFTriDA)	537.1	2.0	---		92.7587	100	ng/L	93	70 - 130	---	---	1.0	01/28/2021 07:57	01/28/2021 23:07	4822025
FBM	Perfluoroundecanoic acid (PFUnA)	537.1	2.0	---		95.4524	100	ng/L	95	70 - 130	---	---	1.0	01/28/2021 07:57	01/28/2021 23:07	4822025
FBM	N-ethyl Perfluorooctanesulfonamidoacetic acid	537.1	2.0	---		92.8729	100	ng/L	93	70 - 130	---	---	1.0	01/28/2021 07:57	01/28/2021 23:07	4822025
FBM	N-methyl Perfluorooctanesulfonamidoacetic acid	537.1	2.0	---		90.9511	100	ng/L	91	70 - 130	---	---	1.0	01/28/2021 07:57	01/28/2021 23:07	4822025
FBM	HFPO-DA/GenX	537.1	2.0	---		97.1346	100	ng/L	97	70 - 130	---	---	1.0	01/28/2021 07:57	01/28/2021 23:07	4822025
FBM	ADONA	537.1	2.0	---		94.7562	100	ng/L	95	70 - 130	---	---	1.0	01/28/2021 07:57	01/28/2021 23:07	4822025
FBM	9C+PF3ONS/F-53B Major	537.1	2.0	---		97.4067	100	ng/L	97	70 - 130	---	---	1.0	01/28/2021 07:57	01/28/2021 23:07	4822025
FBM	11C+PF3OUs/F-53B Minor	537.1	2.0	---		97.3048	100	ng/L	97	70 - 130	---	---	1.0	01/28/2021 07:57	01/28/2021 23:07	4822025
FBM	Perfluorotetradecanoic acid (PFTeDA)	537.1	2.0	---		95.9284	100	ng/L	96	70 - 130	---	---	1.0	01/28/2021 07:57	01/28/2021 23:07	4822025
FBM	SS-HFPO-DA-13C3	537.1	N/A	---		38.3639	40.0	ng/L	96	70 - 130	---	---	1.0	01/28/2021 07:57	01/28/2021 23:07	4822025
FS	Perfluorooctanoic acid (PFOA)	537.1	2.0	Plant 1 Tap-LV Waterworks	<	2.0		ng/L	---	---	---	---	0.96	01/28/2021 07:57	01/29/2021 02:28	4820255
FS	Perfluorooctanesulfonic acid (PFOS)	537.1	2.0	Plant 1 Tap-LV Waterworks	<	2.0		ng/L	---	---	---	---	0.96	01/28/2021 07:57	01/29/2021 02:28	4820255
FS	IS-NMeFOSAA-d3	537.1	N/A	Plant 1 Tap-LV Waterworks		1204.100	1185846.26	ng/L	102	50 - 150	---	---	0.96	01/28/2021 07:57	01/29/2021 02:28	4820255
FS	IS-PFOA-13C2	537.1	N/A	Plant 1 Tap-LV Waterworks		1570511	1522480.84	ng/L	103	50 - 150	---	---	0.96	01/28/2021 07:57	01/29/2021 02:28	4820255
FS	IS-PFOA-13C4	537.1	N/A	Plant 1 Tap-LV Waterworks		6862525	3759097.03	ng/L	102	50 - 150	---	---	0.96	01/28/2021 07:57	01/29/2021 02:28	4820255
FS	SS-NEFOSAA-d5	537.1	N/A	Plant 1 Tap-LV Waterworks		138.9908	160	ng/L	90	70 - 130	---	---	0.96	01/28/2021 07:57	01/29/2021 02:28	4820255
FS	SS-PFDA-13C2	537.1	N/A	Plant 1 Tap-LV Waterworks		35.0501	40.0	ng/L	91	70 - 130	---	---	0.96	01/28/2021 07:57	01/29/2021 02:28	4820255
FS	SS-PFHXA-13C2	537.1	N/A	Plant 1 Tap-LV Waterworks		38.0576	40.0	ng/L	99	70 - 130	---	---	0.96	01/28/2021 07:57	01/29/2021 02:28	4820255
FS	Perfluorobutanesulfonic acid (PFBS)	537.1	2.0	Plant 1 Tap-LV Waterworks	<	2.0		ng/L	---	---	---	---	0.96	01/28/2021 07:57	01/29/2021 02:28	4820255
FS	Perfluoroheptanoic acid (PFHpA)	537.1	2.0	Plant 1 Tap-LV Waterworks	<	2.0		ng/L	---	---	---	---	0.96	01/28/2021 07:57	01/29/2021 02:28	4820255
FS	Perfluorohexanesulfonic acid (PFHxS)	537.1	2.0	Plant 1 Tap-LV Waterworks	<	2.0		ng/L	---	---	---	---	0.96	01/28/2021 07:57	01/29/2021 02:28	4820255
FS	Perfluorononanoic acid (PFNA)	537.1	2.0	Plant 1 Tap-LV Waterworks	<	2.0		ng/L	---	---	---	---	0.96	01/28/2021 07:57	01/29/2021 02:28	4820255
FS	Perfluorodecanoic acid (PFDA)	537.1	2.0	Plant 1 Tap-LV Waterworks	<	2.0		ng/L	---	---	---	---	0.96	01/28/2021 07:57	01/29/2021 02:28	4820255
FS	Perfluorohexanoic acid (PFHxA)	537.1	2.0	Plant 1 Tap-LV Waterworks	<	2.0		ng/L	---	---	---	---	0.96	01/28/2021 07:57	01/29/2021 02:28	4820255
FS	Perfluorododecanoic acid (PFDoA)	537.1	2.0	Plant 1 Tap-LV Waterworks	<	2.0		ng/L	---	---	---	---	0.96	01/28/2021 07:57	01/29/2021 02:28	4820255
FS	Perfluorotridecanoic acid (PFTriDA)	537.1	2.0	Plant 1 Tap-LV Waterworks	<	2.0		ng/L	---	---	---	---	0.96	01/28/2021 07:57	01/29/2021 02:28	4820255
FS	Perfluoroundecanoic acid (PFUnA)	537.1	2.0	Plant 1 Tap-LV Waterworks	<	2.0		ng/L	---	---	---	---	0.96	01/28/2021 07:57	01/29/2021 02:28	4820255
FS	N-ethyl Perfluorooctanesulfonamidoacetic acid	537.1	2.0	Plant 1 Tap-LV Waterworks	<	2.0		ng/L	---	---	---	---	0.96	01/28/2021 07:57	01/29/2021 02:28	4820255
FS	N-methyl Perfluorooctanesulfonamidoacetic acid	537.1	2.0	Plant 1 Tap-LV Waterworks	<	2.0		ng/L	---	---	---	---	0.96	01/28/2021 07:57	01/29/2021 02:28	4820255
FS	HFPO-DA/GenX	537.1	2.0	Plant 1 Tap-LV Waterworks	<	2.0		ng/L	---	---	---	---	0.96	01/28/2021 07:57	01/29/2021 02:28	4820255
FS	ADONA	537.1	2.0	Plant 1 Tap-LV Waterworks	<	2.0		ng/L	---	---	---	---	0.96	01/28/2021 07:57	01/29/2021 02:28	4820255
FS	9C+PF3ONS/F-53B Major	537.1	2.0	Plant 1 Tap-LV Waterworks	<	2.0		ng/L	---	---	---	---	0.96	01/28/2021 07:57	01/29/2021 02:28	4820255

QC Summary Report (cont.)

Sample Type	Analyte	Method	MRL	Client ID	Result Flag	Amount	Target	Units	% Recovery	Recovery Limits	RPD	RPD Limit	Dil Factor	Extracted	Analyzed	EEA ID #
FS	11Cl-PF3OUdSf-53B Minor	537.1	2.0	Plant 1 Tap-LV Waterworks	<	2.0		ng/L	---	---	---	---	0.96	01/28/2021 07:57	01/29/2021 02:28	4820255
FS	Perfluorotetradecanoic acid (PFTeDA)	537.1	2.0	Plant 1 Tap-LV Waterworks	<	2.0		ng/L	---	---	---	---	0.96	01/28/2021 07:57	01/29/2021 02:28	4820255
FS	SS-HFO-DA-13C3	537.1	N/A	Plant 1 Tap-LV Waterworks		36.1941	40.0	ng/L	94	70 - 130	---	---	0.96	01/28/2021 07:57	01/29/2021 02:28	4820255
FS	Perfluorooctanoic acid (PFOA)	537.1	2.0	Plant 2 Tap-LV Waterworks	<	2.0		ng/L	---	---	---	---	0.94	01/28/2021 07:57	01/29/2021 02:39	4820256
FS	Perfluorooctanesulfonic acid (PFOS)	537.1	2.0	Plant 2 Tap-LV Waterworks	<	2.0		ng/L	---	---	---	---	0.94	01/28/2021 07:57	01/29/2021 02:39	4820256
FS	IS-NMeFOSAA-d3	537.1	N/A	Plant 2 Tap-LV Waterworks		1226433	1185846.2E	ng/L	103	50 - 150	---	---	0.94	01/28/2021 07:57	01/29/2021 02:39	4820256
FS	IS-PFOA-13C2	537.1	N/A	Plant 2 Tap-LV Waterworks		1504905	1522480.84	ng/L	99	50 - 150	---	---	0.94	01/28/2021 07:57	01/29/2021 02:39	4820256
FS	IS-PFOS-13C4	537.1	N/A	Plant 2 Tap-LV Waterworks		6616756	3759097.03	ng/L	98	50 - 150	---	---	0.94	01/28/2021 07:57	01/29/2021 02:39	4820256
FS	SS-NEIFOSAA-d5	537.1	N/A	Plant 2 Tap-LV Waterworks		136.1883	160	ng/L	91	70 - 130	---	---	0.94	01/28/2021 07:57	01/29/2021 02:39	4820256
FS	SS-PFDA-13C2	537.1	N/A	Plant 2 Tap-LV Waterworks		36.8381	40.0	ng/L	98	70 - 130	---	---	0.94	01/28/2021 07:57	01/29/2021 02:39	4820256
FS	SS-PFHXA-13C2	537.1	N/A	Plant 2 Tap-LV Waterworks		37.0582	40.0	ng/L	99	70 - 130	---	---	0.94	01/28/2021 07:57	01/29/2021 02:39	4820256
FS	Perfluorobutanesulfonic acid (PFBS)	537.1	2.0	Plant 2 Tap-LV Waterworks	<	2.0		ng/L	---	---	---	---	0.94	01/28/2021 07:57	01/29/2021 02:39	4820256
FS	Perfluorohexanoic acid (PFHxA)	537.1	2.0	Plant 2 Tap-LV Waterworks	<	2.0		ng/L	---	---	---	---	0.94	01/28/2021 07:57	01/29/2021 02:39	4820256
FS	Perfluorohexanesulfonic acid (PFHxS)	537.1	2.0	Plant 2 Tap-LV Waterworks	<	2.0		ng/L	---	---	---	---	0.94	01/28/2021 07:57	01/29/2021 02:39	4820256
FS	Perfluorononanoic acid (PFNA)	537.1	2.0	Plant 2 Tap-LV Waterworks	<	2.0		ng/L	---	---	---	---	0.94	01/28/2021 07:57	01/29/2021 02:39	4820256
FS	Perfluorodecanoic acid (PFDA)	537.1	2.0	Plant 2 Tap-LV Waterworks	<	2.0		ng/L	---	---	---	---	0.94	01/28/2021 07:57	01/29/2021 02:39	4820256
FS	Perfluorohexanoic acid (PFHxA)	537.1	2.0	Plant 2 Tap-LV Waterworks	<	2.0		ng/L	---	---	---	---	0.94	01/28/2021 07:57	01/29/2021 02:39	4820256
FS	Perfluorohexanesulfonic acid (PFHxS)	537.1	2.0	Plant 2 Tap-LV Waterworks	<	2.0		ng/L	---	---	---	---	0.94	01/28/2021 07:57	01/29/2021 02:39	4820256
FS	Perfluorododecanoic acid (PFDDA)	537.1	2.0	Plant 2 Tap-LV Waterworks	<	2.0		ng/L	---	---	---	---	0.94	01/28/2021 07:57	01/29/2021 02:39	4820256
FS	Perfluorotridecanoic acid (PFTrDA)	537.1	2.0	Plant 2 Tap-LV Waterworks	<	2.0		ng/L	---	---	---	---	0.94	01/28/2021 07:57	01/29/2021 02:39	4820256
FS	Perfluoroundecanoic acid (PFUnA)	537.1	2.0	Plant 2 Tap-LV Waterworks	<	2.0		ng/L	---	---	---	---	0.94	01/28/2021 07:57	01/29/2021 02:39	4820256
FS	N-methyl Perfluorooctanesulfonamidoacetic acid	537.1	2.0	Plant 2 Tap-LV Waterworks	<	2.0		ng/L	---	---	---	---	0.94	01/28/2021 07:57	01/29/2021 02:39	4820256
FS	N-methyl Perfluorooctanesulfonamidoacetic acid	537.1	2.0	Plant 2 Tap-LV Waterworks	<	2.0		ng/L	---	---	---	---	0.94	01/28/2021 07:57	01/29/2021 02:39	4820256
FS	HFO-DA/GenX	537.1	2.0	Plant 2 Tap-LV Waterworks	<	2.0		ng/L	---	---	---	---	0.94	01/28/2021 07:57	01/29/2021 02:39	4820256
FS	ADONA	537.1	2.0	Plant 2 Tap-LV Waterworks	<	2.0		ng/L	---	---	---	---	0.94	01/28/2021 07:57	01/29/2021 02:39	4820256
FS	9Cl-PF3ONSf-53B Major	537.1	2.0	Plant 2 Tap-LV Waterworks	<	2.0		ng/L	---	---	---	---	0.94	01/28/2021 07:57	01/29/2021 02:39	4820256
FS	11Cl-PF3OUdSf-53B Minor	537.1	2.0	Plant 2 Tap-LV Waterworks	<	2.0		ng/L	---	---	---	---	0.94	01/28/2021 07:57	01/29/2021 02:39	4820256
FS	Perfluorotetradecanoic acid (PFTeDA)	537.1	2.0	Plant 2 Tap-LV Waterworks	<	2.0		ng/L	---	---	---	---	0.94	01/28/2021 07:57	01/29/2021 02:39	4820256
FS	SS-HFO-DA-13C3	537.1	N/A	Plant 2 Tap-LV Waterworks		37.0981	40.0	ng/L	99	70 - 130	---	---	0.94	01/28/2021 07:57	01/29/2021 02:39	4820256
CCM	Perfluorooctanoic acid (PFOA)	537.1	2.0	---		100.4205	100	ng/L	100	70 - 130	---	---	1.0	01/27/2021 13:34	01/29/2021 03:43	4822377
CCM	Perfluorooctanesulfonic acid (PFOS)	537.1	2.0	---		97.3848	100	ng/L	97	70 - 130	---	---	1.0	01/27/2021 13:34	01/29/2021 03:43	4822377
CCM	IS-NMeFOSAA-d3	537.1	N/A	---		1182060	1182059.73	ng/L	100	50 - 150	---	---	1.0	01/27/2021 13:34	01/29/2021 03:43	4822377
CCM	IS-PFOA-13C2	537.1	N/A	---		1498686	1498686.32	ng/L	100	50 - 150	---	---	1.0	01/27/2021 13:34	01/29/2021 03:43	4822377
CCM	IS-PFOS-13C4	537.1	N/A	---		6535880	3535880.32	ng/L	100	50 - 150	---	---	1.0	01/27/2021 13:34	01/29/2021 03:43	4822377
CCM	SS-NEIFOSAA-d5	537.1	N/A	---		164.6722	160	ng/L	103	70 - 130	---	---	1.0	01/27/2021 13:34	01/29/2021 03:43	4822377
CCM	SS-PFDA-13C2	537.1	N/A	---		39.6399	40.0	ng/L	99	70 - 130	---	---	1.0	01/27/2021 13:34	01/29/2021 03:43	4822377
CCM	SS-PFHXA-13C2	537.1	N/A	---		39.3166	40.0	ng/L	98	70 - 130	---	---	1.0	01/27/2021 13:34	01/29/2021 03:43	4822377
CCM	Perfluorobutanesulfonic acid (PFBS)	537.1	2.0	---		99.6495	100	ng/L	100	70 - 130	---	---	1.0	01/27/2021 13:34	01/29/2021 03:43	4822377
CCM	Perfluorohexanoic acid (PFHxA)	537.1	2.0	---		103.9934	100	ng/L	104	70 - 130	---	---	1.0	01/27/2021 13:34	01/29/2021 03:43	4822377
CCM	Perfluorohexanesulfonic acid (PFHxS)	537.1	2.0	---		96.7141	100	ng/L	97	70 - 130	---	---	1.0	01/27/2021 13:34	01/29/2021 03:43	4822377
CCM	Perfluorononanoic acid (PFNA)	537.1	2.0	---		101.9121	100	ng/L	102	70 - 130	---	---	1.0	01/27/2021 13:34	01/29/2021 03:43	4822377
CCM	Perfluorodecanoic acid (PFDA)	537.1	2.0	---		102.3983	100	ng/L	102	70 - 130	---	---	1.0	01/27/2021 13:34	01/29/2021 03:43	4822377

QC Summary Report (cont.)

Sample Type	Analyte	Method	MRL	Client ID	Result Flag	Amount	Target	Units	% Recovery	Recovery Limits	RPD	RPD Limit	Dil Factor	Extracted	Analyzed	EEA ID #
CCM	Perfluorohexanoic acid (PFHxA)	537.1	2.0	---		96.8053	100	ng/L	97	70 - 130	----	----	1.0	01/27/2021 13:34	01/29/2021 03:43	4822377
CCM	Perfluorododecanoic acid (PFDoA)	537.1	2.0	---		98.4588	100	ng/L	98	70 - 130	----	----	1.0	01/27/2021 13:34	01/29/2021 03:43	4822377
CCM	Perfluorotridecanoic acid (PFTrDA)	537.1	2.0	---		98.3730	100	ng/L	98	70 - 130	----	----	1.0	01/27/2021 13:34	01/29/2021 03:43	4822377
CCM	Perfluoroundecanoic acid (PFUnA)	537.1	2.0	---		105.1932	100	ng/L	105	70 - 130	----	----	1.0	01/27/2021 13:34	01/29/2021 03:43	4822377
CCM	N-methyl Perfluorooctanesulfonamidoacetic acid	537.1	2.0	---		99.2185	100	ng/L	99	70 - 130	----	----	1.0	01/27/2021 13:34	01/29/2021 03:43	4822377
CCM	N-methyl Perfluorooctanesulfonamidoacetic acid	537.1	2.0	---		99.8864	100	ng/L	100	70 - 130	----	----	1.0	01/27/2021 13:34	01/29/2021 03:43	4822377
CCM	HFPO-DA/GenX	537.1	2.0	---		98.9350	100	ng/L	99	70 - 130	----	----	1.0	01/27/2021 13:34	01/29/2021 03:43	4822377
CCM	ADONA	537.1	2.0	---		102.4424	100	ng/L	102	70 - 130	----	----	1.0	01/27/2021 13:34	01/29/2021 03:43	4822377
CCM	9CI+PF3ONS/F-53B Major	537.1	2.0	---		100.9643	100	ng/L	101	70 - 130	----	----	1.0	01/27/2021 13:34	01/29/2021 03:43	4822377
CCM	11CI+PF3OUdS/F-53B Minor	537.1	2.0	---		101.6596	100	ng/L	102	70 - 130	----	----	1.0	01/27/2021 13:34	01/29/2021 03:43	4822377
CCM	Perfluorotetradecanoic acid (PFTeDA)	537.1	2.0	---		107.1008	100	ng/L	107	70 - 130	----	----	1.0	01/27/2021 13:34	01/29/2021 03:43	4822377
CCM	SS-HFPO-DA-13C3	537.1	N/A	---		39.8523	40.0	ng/L	100	70 - 130	----	----	1.0	01/27/2021 13:34	01/29/2021 03:43	4822377

Sample Type Key

<u>Type (Abbr.)</u>	<u>Sample Type</u>	<u>Type (Abbr.)</u>	<u>Sample Type</u>
CCH	Continuing Calibration High		
CCL	Continuing Calibration Low		
CCM	Continuing Calibration Mid		
FS	Field Sample		
FBH	Fortified Blank High		
FBL	Fortified Blank Low		
FBM	Fortified Blank Mid		
LRB	Laboratory Reagent Blank		

END OF REPORT